

Table 1: Refers to “Corrigendum” to be incorporated in revised NIT after 2nd pre-bid meeting.

Sl.	Clause / para no. at	Existing clause/ condition as per NIT	Now to be read as under
1.	Page 1	Date & time for submission & opening of tender document	Date & time for submission & opening of bidding documents
2.	Page 1 & 2	Rate Contract for Supply of Diagnostic Reagents & Chemicals and Consumables against the installation of maintenance free equipment at no cost basis FOR <u>Creation & Establishment of 24 Hours Lab</u> at AIIMS, Rishikesh	Rate Contract for Supply of Diagnostic Reagents & Chemicals and Consumables against installation of maintenance free equipment at no cost basis for <u>establishment of Emergency Lab for a period of five years on fixed terms rate contract basis</u> at AIIMS, Rishikesh.
3.	Page 2	1.(a).All India Institute of Medical Sciences, Rishikesh (AIIMS Rishikesh) invites the offer in two bid system from reputed, Manufacturers/Authorized Dealers/Importers for Rate Contract & Supply of Reagents& Chemicals and Consumables against the installation of maintenance free equipment at no cost basis for all routine parameters to create and establish <u>24 hours lab</u> .	1. (a).All India Institute of Medical Sciences, Rishikesh (AIIMS Rishikesh) invites the offer in two bid system from reputed,Manufacturers/Authorized Dealers/Importers for Rate Contract & Supply of Reagents& Chemicals and Consumables against the installation of maintenance free equipment at no cost basis for all routine parameters to create and establish Emergency lab .
4.	Page 2	(c). AIIMS, Rishikesh will <u>only pay for</u> cost of reagents& chemical or consumables.	(c). AIIMS, Rishikesh will pay only for the cost of reagents& chemicals or consumables.
5.	Page 2	Uptime Guarantee <u>99.5 %</u> of 365 days including all Govt. & hospital holidays.	Uptime guarantee 97 % of 365 days including all Govt. & hospital holidays i.e it should not be more than 24 hours in a month (Breakdown) and a backup for the machine should be available all the time as mentioned in tender.

6.	Page 3	Tender No: 24/diagnostic lab/246/2017-Rish (admn) " <u>before</u> 15.00 hrs on 15/09/ 2017. Bids received after due date shall not be entertained under any circumstances whatsoever. Institute will not be responsible in case of postal delay	Tender No: 24/diagnostic lab/246/2017-Rish(admn) " <u>on or before</u> 15.00 hrs on 20/09/ 2017. Bids received after due date shall not be entertained under any circumstances whatsoever. Institute will not be responsible in case of postal delay
7.	Page 3	6. Bids will be opened on 15/09/2017 at <u>15:30 noon</u> in presence of bidders or their authorized representatives who wish to participate in opening of bid. If opening date happens to be a closed day/holiday, tender will be opened on next working day at same time.	6. Bids will be opened on 20/09/2017 at <u>15:30 hrs</u> in presence of bidders or their authorized representatives who wish to participate in opening of bid. If opening date happens to be a closed day/holiday, Bids will be opened on next working day at same time.
8.	Page 3	7. Any further clarification and / or corrigendum(s) shall be communicated by Tender Office, through website <u>www.aiimsrishikesh.edu.in.</u>	7. Any further clarification and / or corrigendum shall be communicated by Tender Office, through website <u>www.aiimsrishikesh.edu.in. and shall be displayed at notice board</u>
9.	Page 4	Sub:Tender No.: 24/diagnostic lab/246/2017-Rish(admn) 4. I/We hereby agree to abide by all terms and conditions, stipulated by <u>AIIMS Rishikesh</u> in connection with delivery, penalty etc. Rates are being submitted under separate covers, and sheets and shall be considered on their face value.	Sub:Tender No.: 24/diagnostic lab/246/2017-Rish(admn) due on or before 20/09/2017 upto 15 hrs. 4. I/We hereby agree to abide by all terms and conditions <u>as stipulated in tender document</u> in connection with delivery, penalty etc. Rates are being submitted under separate covers, and sheets and shall be considered on their face value.
10.	Page 5	A.: To be filled by bidder and submitted along with Technical Bid.	<u>A.: To be filled by Bidder and submitted along with Technical Bid (PART I)</u>
11.	Page 6	13. Last three years audited balance sheet & profit loss statement of your firm duly certified by CA	13. Last three financial years audited balance sheet & profit loss statement of your firm duly certified by CA

12.	Page 6	<u>B: To be filled by Bidder and submitted along with Price Bid</u>	<u>B: To be filled by Bidder and submitted along with Price Bid (PART II)</u>
13.	Page 8	iv) Eligible Goods and Services: - All goods and related services to be supplied under contract shall have their origin in India or any other country of origin with which India has not banned trade relations. term “origin” used in this clause means place where goods are mined, grown, produced, or manufactured or from where related services are arranged and supplied.	iv) Eligible Goods and Services: - All goods and related services to be supplied under contract shall have their origin in India or any other country of origin with which India has not banned trade relations. Term “ <u>country of origin</u> ” used in this clause means; place where goods are mined, grown, produced, or manufactured or from where related services are arranged and supplied.
14.	Page 10	8. Tender Prices: - Offer/Bid will be submitted in the prescribed format only.	8. Tender Prices: - Offer/Bid will be submitted in the prescribed format in <u>Indian Currency</u> only.
15.	Page 15	<u>Reagents should be IVD/FDA approved and of good quality approved by technical committee of AIIMS Rishikesh</u>	<u>Reagents should have US FDA / Self Declared IVD/ Certificate of Analysis</u> and of good quality approved by technical committee of AIIMS, Rishikesh
16.	Page 21	3. Bidder must provide the make of reagent in Technical as well as Financial bid.	3. Bidder must provide the make of reagent in Technical as well as Financial bid. Bidder must provide HSN code under GST rules of all reagents in Technical as well as Financial bid.
17.	Page 15 & 21	Elisa Reader	Elisa Reader & Washer & should be US FDA approved.
18.	Page 21	Fully Auto Bio Chemistry Analyzer with Electrolyte - <u>Having a high through put US FDA approved & State of Art Quality</u>	<u>Analyzer having a through put of 400 test/hour or more for photometric with ISE US FDA approved & State of Art Quality</u>
<i>However other terms & conditions of NIT will remain same.</i>			

Table 2: Refers to “Addendums” to be incorporated in revised NIT after 2nd pre-bid meeting.

Sl.	To be added at page of NIT	Addendums
6	At page 12 in clause no. 21 as 21 (c)	If any make or model of existing system discontinued in future, the firm may provide similar equipment of other brand but per test cost will remain same and no price hike is permissible of reagents & chemicals till validity of rate contract. Frequency of calibration of all the equipment which will be supplied must be at least quarterly.