

## Corrigendum

**In tender document  
Tender Enquiry No.24/Micro/131(III)/2016-RISH (ADMN)**

Dated: 05-08-2016

As per schedule, Pre- Bid meeting of "Automated Bacterial Identification & Susceptibility System with Accessories for Department of Microbiology" was held on 22-07-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Micro/131(III)/2016-RISH (ADMN) has been made.

<b>AUTOMATED BACTERIAL IDENTIFICATION AND SUSCEPTIBILITY SYSTEM</b>		
<b>Specification point no.</b>	<b>Tender Specification</b>	<b>To be replaced with</b>
2.	The system must have the capacity to accommodate a minimum of 50 tests (either 50 ID and/or AST tests), at any time.	The system must have the capacity to accommodate a minimum of 50 tests <b>(50 ID and 50 AST tests)</b> , at any time.
3.	The system must have a bar code scanning device for test card identification and specimen number entry	The system must have a bar code scanning device for test card identification and specimen number entry
4.	The system must have identification and antibiotic susceptibility testing cards/panels for Gram negative, Gram positive and Yeast and identification for Neisseria and Haemophilus	The system must have identification and antibiotic susceptibility testing cards/panels for Gram negative, Gram positive and Yeast and identification for Neisseria and Haemophilus
7.	The system must have separate cards for Identification and susceptibility testing.	The system must have separate cards/panels for Identification and susceptibility testing.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 23-08-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.