

Corrigendum

**In tender document
Tender Enquiry No.24/Micro/131(VII)/2016-RISH (ADMN)**

Dated: 05-08-2016

As per schedule, Pre- Bid meeting of "Tender for Anaerobic Workstation with Accessories for Department of Microbiology" was held on 22-07-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Micro/131(VII)/2016-RISH (ADMN) has been made.

ANAEROBIC WORK STATION			
Specification point no.	Tender Specification		To be replaced with
1.	Capacity(Petri Dishes)	350-450	Minimum 450
2.	Port/Airlock Capacity	10-12 plates per port	Minimum 10-12 plates per port
3.	Porthole System	Manual/semi-automatic	semi-automatic
5.	Footswitch	Hardwired/wireless	wireless
11.	Refrigeration	Preferable	5 deg C above ambient to 45 deg C
22	A UV light lamp to be provided for disinfecting the chamber		A UV light lamp to be provided in the chamber

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 23-08-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.