

Corrigendum

In tender document Tender Enquiry No.24/OT/121(I)/2016-RISH (ADMN)

Dated: 30-06-2016

As per schedule, Pre- Bid meeting of "Tender for Heavy Duty Surgical Suction Machine for Operation Theatre" was held on 27-06-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/OT/121(I)/2016-RISH (ADMN) has been made.

Technical Specification at page no. 19:-

Following specification is being added.

1. Firm should be USFDA / European CE approved.
2. Noise level should be less than 42 dB.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 19-07-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.