

Corrigendum

Modification in General/Technical specifications:

High End Cardiac HDU/CCU Bed:

As per schedule, Pre- Bid meeting of "Tender for High End Cardiac HDU Monitors for Department of Cardiology." was held on 12-10-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/HDU/258/2017-Rish(Admn) has been made.

- ❖ **Name of tender to be read as Tender for High End Cardiac HDU Beds for Department of Cardiology in Place of Tender for High End Cardiac HDU Monitors for Department of Cardiology.**

Point Srl no. 02 at page no. 18:-

For:-

Dimensions of bed	
Length x Wide	: 2280 mm x 950 mm or better
Mattress Size	: 840 x 2060 mm
Permissible load capacity	: 250 kg or more

Read as:-

Dimensions of bed	
Length x Wide	: 2060-2080 mm x 840-860 mm or better
Mattress Size	: 840 x 2060 mm
Permissible load capacity	: 250 kg or more

Point Srl no. 03 at page no. 18:-

Should have steeples adjustment for the following: -

Height	:	min. 485 mm or less (electrically)
		max. 880 mm or better (electrically)
Back section	:	0- 70 degrees or better (electrically)
Thigh Section	:	0 - 40 degrees or better (electrically)
Shank Rest Adjustment	:	20 - 22 degrees or better (pneumatically)
Trendelenburg	:	15 degrees or better (electrically)
Anti-Trendelenburg	:	13 degrees or better (Electrically)
Lateral Tilt	:	0- 25 degrees or better (Electrically)

Cardio logical chair, fowler position, electric CPR, examination position, anti-shock position
Electrically operated by using single button on central controller.

Back Rest Segment Auto regress: 120 mm

Thigh Rest Segment Auto regress: 90 mm

Read as:

Should have steeples adjustment for the following: -

Height	:	min. 485 mm or less (electrically)
		max. 880 mm or better (electrically)
Back section	:	0- 70 degrees or better (electrically)
Thigh Section	:	0 - 40 degrees or better (electrically)
Shank Rest Adjustment	:	20 - 22 degrees or better (pneumatically)
Trendelenburg	:	15 degrees or better (electrically)
Anti-Trendelenburg	:	13 degrees or better (Electrically)
Lateral Tilt	:	0- 20 ± 2 degrees or better (Electrically)

Cardio logical chair, fowler position, electric CPR, examination position, anti-shock position
Electrically operated by using single button on central controller.

Back Rest Segment Auto regress: 100 - 120 mm

Thigh Rest Segment Auto regress: 70-90 mm

Point Srl no. 09 at page no. 18:-

For:- It should have the facility of lateral tilt adjustment to allow conducting positional therapy, for comfort and safety of patients with injuries of Hip, Spine etc & for convenience of staff in patient handling.

Read as:- It should have the facility of lateral tilt adjustment to allow conducting positional therapy, for comfort and safety of patients with injuries of Hip, Spine etc & for convenience of staff in patient handling, **preferably automatic.**

Point Srl no. 10 at page no. 18:-

For:-It should have facility of bed length extension for large patients up to 20 cm. Bed should have inbuilt retractable linen shelf.

Read as:- It should have facility of bed length extension for large patients up to 20 ± 1 cm. Bed should have inbuilt retractable linen shelf

Point Srl no. 13 at page no. 18:-

For:-It should be equipped with a central control panel with facility of locking the functions to prevent accidental change of position by patient / attendant, large LCD display to show bed positions.

Read as: It should be equipped with a central control panel with facility of locking the functions to prevent accidental change of position by patient / attendant, large LCD 6-4" display to show bed positions.

Point Srl no. 20 at page no. 18:-

For:- Mattress of the Bed should be made up of High quality medical grade foam with Anti-Microbial agent incorporated into all components for Prohibiting growth of bacteria & fungi. Mattress catalogue should also be attached with certifications for the mattress material.

Read as:- Mattress of the Bed should be made up of High quality medical grade foam with Anti-Microbial agent incorporated into all components for Prohibiting growth of bacteria & fungi. Mattress catalogue should also be attached with certifications for the mattress material. Mattress should be from same manufacturer.

Point Srl no. 21 at page no. 18:-

For:- Base frame should be covered with plastic shrouds for ease of cleaning and disinfection.

Read as:- IP 54/IP 66 or better.

Point Srl no. 23 at page no. 19:-

For :

Should comply with international safety and quality standards for medical equipment –European CE, MDD/93/42/EEC, ISO 13485 and ISO 14001.

Read as:-

Should comply with international safety and quality standards for medical equipment –European CE/ USFDA/ ISO 14001 & ISO 13485.

Point Srl no. 27 at page no. 19:-

For:-

Each bed should be supplied complete with:

- a) Bed Ends: 01 pair
- b) split Side Rails with controls: 01 pair
- c) IV Rod : 01 No.
- d) Mattress: 01 No.
- e) Bed side locker: 01 No.
- f) Over Bed table: 01 No.

Read as:

Each bed should be supplied complete with:

- Mattress & all accessories should be from same manufacturer.
 - a) Bed Ends: 01 pair
 - b) split Side Rails with controls: 01 pair
 - c) IV Rod : 01 No.
 - d) Mattress: 01 No.
 - e) Bed side locker: 01 No.
 - f) Over Bed table: 01 No.

Following point is being Added:

- **The quantity mentioned in the tender documents is tentative. 12 more beds will be increased as per the requirement of Institute.**
- **95% working time**
- **5 years' warranty with 5 year CMC (including spares & all electronic items)**
- **Any Consumables- Please quote the rate of consumables, rate would be frozen for next 5 years.**
- **2 mandatory visit per year by engineer for maintenance of equipment apart from repair call.**

Note: - Last date of submission of bids in respect of aforesaid tender is hereby extended till 15-11-2017 at 03:00 PM and Technical bid will be opened on next day at 03.30 PM.