

ALL INDIA INSTITUTE OF MEDICAL SCIENCE
RISHIKESH

Date-17/12/2014

Report of the Tender Evaluation Committee

The Tender Committee opened the tender of "DVT Pump" Tender Enquiry No. 24/Ortho/64(7)/2014-Rish(Admn) in front of participating bidders on 31/10/2014 and analyzed/scrutinised administratively and technically of all 03 tenders received.

2. After administrative and technical evaluation by the committee they found following 03 bids fulfil eligibility criteria for qualification of technical bids and thus declared technically qualified for further sample evaluation. These firms are as under:-

S/No	Tender Srl No	Firma Name
1	64(7)-1/3	M/s Relief Surgical
2	64(7)-2/3	M/s Adision Equipment Co.
3	64(7)-3/3	M/s NCCF of India Ltd

3. After that Technical Evaluation Committee assessed the samples of 03 technically qualified bidders. As per report of the Technical Committee firm qualified or rejected (reason wise) is as under:-

S/No	Tender Srl No	Firms Name & Address	Item Name	Reason
1	64(7)-1/3	M/s Relief Surgical	DVT Pump	Rejected due to :- Sequential circumferential & gradient compression not present. No visual alarm.
2	64(7)-2/3	M/s Adision Equipment Co.		Qualified
3	64(7)-3/3	M/s NCCF of India Ltd		Qualified

4. The Tender Committee has accordingly opened the financial bid of technically qualified firms and considered their financial bids only in respect of their only those firm declared qualified by Technical Evaluation Committee.

4. Financial bid of aforesaid firms will be open on **19-12-2014 at 11.00 AM** in the tender office, AIIMS Rishikesh. All successful bidders are requested to participate in the opening of Financial Bid.