

Technical Specification for Arthroscopic Limb Positioner Spider 2 (Smith & Nephew)

21

Arthroscopic Limb Positioner:

Specifications

- Should be battery operated limb positioner for knee ,hip, shoulder, wrist & ankle arthroscopies
- Should offer instantaneous limb repositioning with in sterile zone itself
- Should deliver exceptional control of abduction, rotation & forward flexion in beach chair as well as lateral decubitus position in shoulder surgeries
- Should come with a distal activation switch which can be attached to the sterile drape of the patient and can be used to reposition the shoulder by applying small pressure
- Should come with square rail clamps to connect it to the OT table
- Should have the provision for additional battery back up
- Should come with the battery charger
- Should come with two sets of traction accessories for the limb positioner & traction requirements
- Should come with a cart to make it comfortable to move from one place to another

