

SPECIFICATION OF CARDIAC PROBE

REGULAR ADULT & PAEDIATRIC ECHO TRANSDUCER PROBE

- 1. Should be able to perform Echo in both normal size adult & large children.
- 2. Compatible with the available USG machine in department of Anaesthesia, AIIMS, Rishikesh.
- 3. USG Machine Model No. Logic-e (GE)
- 4. Frequency 3-8 Mhz.
- 5. This probe should use either single crystal technology or pure wave technology for excellent image quality on difficult to image patients.
- 6. Must be US FDA approved.
- 7. Demo is must.

Swingh

PL

Dr. Anil Kumar

Dr. Anil Kumar