

# Invitation of Quotation

For

“DVT Pumps for the Dept. of  
CTVS”

At

All India Institute of Medical Sciences, Rishikesh

**Issue Date** : 20 May, 2019  
**Inquiry No.** : F.No.32/177/2019(RIS)  
**Last Date of Submission** : 26 May,2019 at 05:00 PM.



**All India Institute of Medical Sciences, Rishikesh**

Virbhadrā Road, Rishikesh, Uttarakhand-249201.

Telephone: 0135-2462913 , email: [www.aiimsrishikesh.edu.in](http://www.aiimsrishikesh.edu.in)

## **Invitation of Quotation for Department of CTVS at AIIMS Rishikesh.**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Rishikesh for supply of Equipment's for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 26.05.2019, 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under: -

### **"QUOTATION FOR Department of CTVS FILE NO. 32/177/2019 (RIS) DUE ON 26/05/2019, 05.00 PM"**

#### **1. Terms & Conditions:**

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- C) Rates must be quoted as per the format specified Taxes extra if any must be written separately.
- D) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- E) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- F) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.  
**L1 will be decided on individual item basis.**
- G) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- H) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Uttarakhand/Central Govt.
  - The firm shall have valid VAT/ Sales Tax No. and IT PAN.
  - **The firm should not be black listed by any Govt. Agency/Dept.**
- I) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- J) The supplier may be asked to submit a sample of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.
- K) **Delivery Period** – 15 days from issuing purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be made only after satisfactorily delivery, commissioning and inspection of material by the AIIMS Rishikesh.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the Supplier and any other department of AIIMS Rishikesh with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Rishikesh whose decision will be final and binding upon the Supplier.
- O) AIIMS, Rishikesh reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Rishikesh will be final in this regard.
- P) AIIMS, Rishikesh reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Rishikesh will be final in this regard.

**Administrative Officer**

Encl.: Annexure 1 (Format of Price Bid)

[On the letterhead of firm]

**ANNEXURE - I**  
**PRICE BIDFORM**

To,  
Administrative Officer,  
AIIMS, Rishikesh.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. **"QUOTATION FOR Department of CTVS AGAINST THE INQUIRY NO. FILE NO. 32/177/2019(RIS)" DUE ON 26/05/2019, 05.00 PM** for Supply of following items at AIIMS Rishikesh".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S.No	Item Name	Qty	Make	Unit price	Tax %	Total amount
1.	DVT Pumps	02				

Date \_\_\_\_\_

Place \_\_\_\_\_

Signature of Authorized Person) :- \_\_\_\_\_  
(Name) \_\_\_\_\_  
Name of Firm/Company/Agency \_\_\_\_\_  
Phone No. \_\_\_\_\_  
Email: \_\_\_\_\_

### PHNEUMATIC COMPRESSION DEVICE (DVT PUMP) SPECIFICATION

- The System should provide complete solution to meet all Intermittent Pneumatic Compression requirements for prevention of DVT by reducing venous stasis
- Should have Distinctive single tube design to deliver both uniform and sequential modalities from one easy-to-use pump
- The system should be compatible to three types of sleeves i.e. calf length, thigh length and foot.
- When foot sleeve is attached, system should deliver uniform compression pressure of 130mmHg for 3 second in 30 second (compression + decompression) cycle.
- Should have Cable management system to reduce tubing around bed/chair space
- Should have Integrated battery to ensure continuation of therapy during transportation
- Should have Fixed tube sets to be ready for therapy.
- Should have snap lock connectors with Auto Garment Recognition automatically adjusts compression cycle without additional user intervention
- Should have Auto garment recognition on foot, calf and calf & thigh garments
- Mains power on/off button Illuminated when pump is operating.
- Should have Large LCD screen to view and manage therapy status with Large LED light display to ensure visibility from all angles
- Should have Patient Run Hours Meter to view time compression has been delivered
- Should have Real Time Pressure Indication to ensure correct pressure is being delivered
- Should have Integrated carry handle to aid transportation and swing out bed hooks for mounting system to bed with built in power cord storage.
- Should have Clear audible and visual alarm indicators to assure that the pump is delivering the correct pressure at all times. Visual Alarm status is indicated by color change in LED.
- Should have Convenience of single pump for all active compression therapy
- Garments available should be Lightweight, breathable fabric garments ensure comfort and help prevent the build-up of heat and moisture. Garment should be non-thermal in nature with excellent moisture management properties for patient comfort and patient compliance.
- Should be available in following size of garments (MUST):
- CALF SIZE: 43Cms/58Cms/81Cms circumference.
- CALF & THIGH SIZE: 71Cms/89Cms circumference.

UNITS: 02

GARMENTS: 04 pairs. Each size or as per choice.

*Rajesh*