

Invitation of Quotation


For

Fully Automated BP Monitor

At

All India Institute of Medical Sciences, Rishikesh

Issue Date : 17 Jan. 2018
Inquiry No. : F.No.32/449/2018(RIS)
Last Date of Submission : 23 Jan, 2018 at 05:00 PM.



**All India Institute of Medical Sciences,
Rishikesh**

Virbhadrā Road, Rishikesh, Uttarakhand-249203.
Telephone: 0135-2462913 , email: www.aiimsrishikesh.edu.in

Invitation of Quotation for Dept. of Transfusion Medicine

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Rishikesh for supply of Items for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 23.01.2018, 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

" QUOTATION FOR Department of Transfusion Medicine FILE NO. 32/449/2018(RIS) DUE ON 23.01.2018, 05.00 PM"

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
- B) Quotations must be in the enclosed prescribed Proforma on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- C) Rates must be quoted as per the format specified Taxes extra if any must be written separately. The limit of file is under 2.5 Lakh.
- D) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- E) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- F) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
L1 will be decided on individual item basis.
- G) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- H) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Uttarakhand/Central Govt.
- **The firm shall have valid GST and IT PAN.**
- **The firm should not be black listed by any Govt. Agency/Dept.**
- I) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- J) The supplier may be asked to submit a sample of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.
- K) **Delivery Period** – 15 days from issuing purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be made only after satisfactorily delivery, commissioning and inspection of material by the AIIMS Rishikesh.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the Supplier and any other department of AIIMS Rishikesh with regards to the interpretation of "Terms & Conditions" of this inquiry, the same shall be referred to the Director, AIIMS Rishikesh whose decision will be final and binding upon the Supplier.
- O) AIIMS, Rishikesh reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Rishikesh will be final in this regard.
- P) AIIMS, Rishikesh reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Rishikesh will be final in this regard.

Administrative Officer

Encl.: Annexure 1 (Format of Price Bid)

[On the letterhead of firm]

ANNEXURE - I
PRICE BIDFORM

To,
Administrative Officer, AIIMS,
Rishikesh.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR Transfusion Medicine. AGAINST THE INQUIRY NO. FILE NO. 32/449/2018(RIS)" DUE ON 23.01.2018, 05.00 PM for Supply of following items at AIIMS Rishikesh".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Items	Make	Qty.	Unit Price	Tax	Total Amount
1.	Fully Automated BP Monitor		01			

Date _____
Place _____

Signature of Authorized Person) :- _____
(Name) _____
Name of Firm/Company/Agency _____
Phone No. _____
Email: _____

Name of items with full specifications & required accessories	Quantity(in figures and words)
<p>Fully Automated BP Monitor: Should have following specifications:</p> <ol style="list-style-type: none"> 1. Measurement method: Oscillometric 2. Pressurization: Internal micro pump 3. Pressure Display range: 0 to 299 mmHg 4. NIBP measurement range: SYS 40-270mmHg, DIA 20-200mmHg, Pulse 30-240mmHg 5. Measurement Accuracy Pressure: ± 3 mmHg Pulse: $\pm 5\%$ 6. NIBP Clinical test: BS EN1060-4:2004, BHS:1993, ± 5mmHg SD:8mmHg 7. Cuff fastening method: Torque Controlled Belt drive Method (TCBM) 8. Accuracy, Pulse Rate: ± 5 percent 9. Clock display: 24 hour WITH DATE & TIME 10. Display Type: 3-digit display by LED 11. Cuff: Automatic Operation 12. Power: Switching 100-240V AC 50-60Hz 13. Dimensions: 241(W) x 324(H) x 390(D) 14. Weight: Approx. 5.5 Kgs. 15. Clock function: Clock display , Date & Time (2015 - 2050) 16. Printer: Thermal printer with Auto cutter in monitor 17. Certificate: USFDA/ European certification/ CE 18. For any breakdown company 	1 Nos. /One

<p>should repair within 24 hours and should have engineering service in surrounding 100 km.</p> <p>19. Satisfactory user certificate should be submitted from 3 reputed government organizations.</p> <p>20. Warranty of the system for 5 years.</p> <p>21. Quality certification: ISO certified.</p> <p>22. Should provide user manual.</p>			
--	--	--	--

3. For equipment, please provide the following information