



Corrigendum-II

Date- 10/12/2025

Tender enquiry: 24/PMR/RPMSSystem/853 /2025-Rish (Admn)

Tender Name: Procurement of Repetitive Peripheral Magnetic Stimulation System for the Department of PMR

Amendment in Technical Specifications:

<u>REPETITIVE PERIPHERAL MAGNETIC SYSTEM</u>		Date: 25-11-25
<u>FILE NO – D-21019/64/2024-53</u>		<u>Comp No.18905</u>
Technical Specification		Final Specifications
1. Must have a deep penetrating treatment effect by 3 Tesla		1. Must have a deep penetrating treatment effect by 3 Tesla
2. Must have a magnetic frequency adjustable from 1 to 100 Hz		2. Must have a magnetic frequency adjustable from 1 to 100 Hz
3. Must have a non-invasive and non-contact transducer for treatments		3. Must have a non-invasive and non-contact transducer for treatments
4. Must have two types of magnetic transducers for wider application		4. Must have two types of magnetic transducers for wider application
5. Must have an oil circulation cooling transducer for continuous treatment	This is for your kind information that worldwide there is two type of coil cooling systems (Oil Based & Air cooled), so we request you kindly add Air cooling system also. (ACCEPTED)	5. Must have an oil circulation cooling or air cooling transducer for continuous treatment



अखिल भारतीय आयुर्विज्ञान संस्थान ऋषिकेश
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
केंद्रीय भंडार/ Central Store

Email: fic.procurement@aiimsrishikesh.edu.in

6. Must have 3 treatment modes such as Auto, Manual and User modes with 20 protocols in 7.each mode			6. Must have 3 treatment modes such as Auto, Manual and User modes with 20 protocols in 7.each mode
8.Must have a provision to add additional Urinary Incontinence UI chair to provide patients with toilet position for more effective treatment of Urinary Incontinence		Must be supplied with a urinary incontinence (UI) chair to provide the patient with toilet position for more effective treatment of urinary incontinence. (ACCEPTED)	8. Must be supplied with a urinary incontinence (UI) chair to provide the patient with toilet position for more effective treatment of urinary incontinence
9.Must have an 8" touch screen display			9.Must have an 8" touch screen display
10.Must have a large applicator on the handle for comfort of use			10.Must have a large applicator on the handle for comfort of use
11.Must have a small hand-held applicator for the manual development of the treatment area			11.Must have a small hand-held applicator for the manual development of the treatment area
12.Must have an adjustable treatment time from 1 to 60 minutes.			12.Must have an adjustable treatment time from 1 to 60 minutes.
13.Operating power supply must be 230 VAC, 50-60 Hz			13.Operating power supply must be 230 VAC, 50-60 Hz

