



अखिल भारतीय आयुर्विज्ञान संस्थान ऋषिकेश
ALL INDIA INSTITUTE OF MEDICAL SCIENCES

केंद्रीय भंडार/ Central Store

Email: fic.procurement@aiimsrishikesh.edu.in

Corrigendum-II

Date- 10/11/2025

Tender enquiry: 24/Pulmonary/FOT with Spirometry/845/2025-Rish

Tender Name: Procurement of FOT With Spirometry for the Department of Pulmonary Medicine

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, RISHIKESH

DEPARTMENT OF PULMONARY MEDICINE

Technical Specification Corrigendum

In reply to representations received (3) following changes are recommended

Sr. No.	Original Technical Specification	Should be read as
01	Should have combination of both Spirometry and Oscillometry in same device which should work on common software.	Should have combination of both spirometry & oscillometry which works on common software
11	Spirometry System should be based on Pneumotach Flow Measurement technology.	Spirometry system should be based on Pneumotach or ultrasonic flow sensor technology
15 C	Antibacterial Filters : 100 Nos	Antibacterial Filters : 1000 Nos

Delivery time can be modified upto 90 days as per administrative norms /administrative approval