

**Technical Specification for Perineal Puncture Attachment  
Department of Urology**

Sl No	Technical Specification	Demand Qty
1	The perineal puncture attachment consists of a needle guide and a mounting ring with clamp	1
2	The needle guide of perineal puncture attachment should comprises 9 parallel guide channels, spaced 5mm apart, each with an internal diameter of 2.1mm, suitable for a 14-gauge needle.	
3	The guide should be parallel to the centerline of the transducer.	
4	The perineal puncture attachment should be reusable	
5	It should be made of high quality stainless steel	
6	It should be compatible with autoclave/ plasma/ETO sterilization	
7	It should be compatible with the existing Colour Doppler System Prostate triplane Transducer ( Model No- Flex Focus 800, BK Make)	
<b>Other requirements</b>		
1	Should have provided with 2 years warranty	
2	Any Complaint to be attend with in 48-72 hrs	
3	Atleast 3 maintenance visit to be made by service engineer per year in warranty period, this is apart from the repair / breakdown call	
4	The company should have the certification of ISO/ USFDA/BIS/European CE or any international certification whichever is applicable.	
5	Rate freeze for accessories for 2 years	



Klia Whittington  
Vice President, Commercial Operations

BK Medical  
Mileparken 34  
DK-2730 Herlev  
Denmark

O +45 4452 8100  
F +45 4452 8199

gehealthcare.com

22<sup>nd</sup> May 2024

All India Institute of Medical Sciences  
Rishikesh

Reference: Biopsy Guides for BK Medical products

Dear Sir:

BK Medical ApS (a GE HealthCare Company) ("BK Medical") is the proven and reputable manufacturer of BK Medical ultrasound systems. BK Medical's portfolio of products also includes accessory and disposable products for use with certain BK Medical products, such as the following items:

- Puncture Guide (reusable) for 9002 (9C2 High Frequency Curved Array Transducer) - UA1250
- Reusable Dual Biopsy Guide for 9018 - UA1328
- Reusable Perineal Biopsy Guide - UA1324
- Disposable Sterile Dual Biopsy Guide for 9018 - UA1329-S.

BK Medical's commitment to providing top quality products to its end-user customers extends to its accessory and disposable products. As an authorized distributor in India, Healthware is only permitted to sell accessory and disposable products approved and provided by BK Medical.

BK Medical certifies that the above-listed products are approved and provided by BK Medical for use with the applicable BK Medical system(s) and/or transducer(s). BK Medical cannot vouch for the quality or compatibility of any alternate products not provided by BK Medical.

Sincerely,

*Klia Whittington*

Klia Whittington

P-3 FORM

(to be attached with P-2 form for Proprietary items)

AIIMS Rishikesh

PROPRIETARY ARTICLE CERTIFICATE

It is certified that the items (Reusable perineal Biopsy Guide - UA1324) required in the P-2 form should be purchased from M/s. Healthware Pvt. Ltd. To the best of my knowledge M/s. Healthware Pvt. Ltd. are the sole manufacturer/agents of the sole manufacturers M/s. Wipro GE Healthcare Pvt. Ltd.

Similar items manufactured by other firm(s) shall not be suitable for our purpose for the following reasons:-

- ① Use of any other make spares damage the original machines
- ② Reusable perineal Biopsy Guide - UA 1324 of other manufactures not match with existing Ultrasound - BK Medical machines (Model No. - Flex Focus 800)

(Sign of Indenter)

Dated

Designation

Department

Recommendation:

डॉ० अंकुर मित्तल / Dr. Ankur Mittal  
अपर- आचार्य एवं प्रमुख  
Additional Professor & Head  
मूत्ररोग चिकित्सा विभाग / Department of Urology  
एम्स रिकेश / AIIMS Rishikesh

डॉ. दिलीप कुमार सिंह  
Dr. Deelip Kumar Singh  
सहायक आचार्य / Assistant Professor  
मूत्ररोग चिकित्सा विभाग  
Department of Urology  
एम्स रिकेश / AIIMS, Rishikesh

Signature of Head of Department/Section

डॉ० विकास कुमार पंवार  
Dr. Vikas Kumar Panwar  
सह-आचार्य / Associate Professor  
मूत्ररोग चिकित्सा विभाग  
Department of Urology  
एम्स रिकेश / AIIMS, Rishikesh

N.B.: The indenter before recording the above certificate should satisfy himself that the article is genuinely of proprietary nature manufactured under patent laws.