

Subject: Reply of representation regarding tender specifications at pre-bid meeting held on 30-03-2026.

Bid No.: GEM/2026/B/7386428 Dated 24-03-2026, Inverted Microscope

Specification with serial number mentioned as per the ATC attached	Suggested amendment to make the specification general	Remarks/justification for amendment suggested	Accepted/ Not accepted
19. Microscope objectives, Camera and licensed software should be from same manufacturer.	Microscope objectives, Camera and licensed/non licensed software should be from same manufacturer.		Not Accepted
24. The item should be certified by BIS/ ISO/ USFDA/ European CE or equivalent.	24. The item should be certified by BIS/ ISO/ USFDA/ European CE/CDSCO or equivalent.	As per recent notification issued by CDSCO. All procurement agencies including hospital, health institutes are requested to make mandatory Manufacturing license of the quoted product/ model from CDSCO as a part of the requirement for procurement of medical device. Without which the medical device cannot be sold in the country.	Accepted


8-4-2026
Indentor
Conc. Department


H.O.D.
Conc. Department

Head of Department
Transfusion Medicine
AIIMS, Rishikesh