

# Corrigendum

**02-11-2024**

As per the scheduled programme, the pre-bid meeting for “Tender for Automated ESR Analyser on Cost per Test Basis” was held on 17-10-2024 at 03.00 PM in the Tender Opening Room. One representation was received in the meeting which was sent to the Indenter Faculty for consideration.

After consideration by indenter faculty following amendment in Tender Enquiry 24/RC/ESR Analyser/760/2024-RISH (ADMN) have been made.

## **Point srl no 01 at page 26.**

**For:** The working principal of the equipment should be based on red cells aggregation and the results correlated with westergren method.

**Read as:** The working principal of the equipment should be based on red cells aggregation/ Sedimentation and the results correlated with westergren method.

## **Point srl no 07 at page 26.**

**For:** Sample results should be provided in less than half a minute.

**Read as:** Sample results should be provided within 30 minutes.