

PROPRIETARY CERTIFICATE

To Whomsoever It May Concern

We, **GymnaUniphy N.V.**, are the sole manufacturer of **Myofeedback and Combined therapy device**, Model: **Gymna Myo 200** and it is further certified that **Gymna Myo 200** is our Proprietary Design.

It is also confirmed that the said product with the given specifications is not manufactured by any other manufacturer in the world.

Specification:

- 1) The unit has possibility to combine electro stimulation with EMG or pressure feedback.
- 2) Target capture function independently for both channels.
- 3) Possibility for the registration of maximum, minimum and average measuring results.
- 4) Possibility to increase resolution around captured target during measurement process (on both channels).
- 5) Audio signal (adjustable) upon achieved the target value.
- 6) Stimulation and relaxation mode function.
- 7) The unit has one pressure feedback in approx range of 0-1000 mmHg for absolute input and 0-750 mmHg for relative input.
- 8) The unit has Pelvic re-education therapy & Muscle re-education therapy.
- 9) The unit has two independent electrotherapy channels for rehabilitation, muscle strengthening, pain management, Unidirectional currents.
- 10) The unit has contra-indications list.
- 11) Capacitive touch screen Monochrome LCD with backlight.
- 12) The unit has Protocols: Objectives & Indications list.
- 13) The unit has more than 80 treatment indications.
- 14) Facility for treatment by setting objectives
- 15) Facility of free memory slots to store programs.
- 16) Optional vacuum unit with continuous and pulsed mode for treatment by vacuum electrodes for electrotherapy.
- 17) Vacuum pressure up to approx 35-300 hPa for continuous mode.
- 18) Vacuum pressure up to of 55- 450 hPa for Pulsation mode.
- 19) Supplied with complete range of probes for EMG feedback and electro stimulation and pressure feedback (anal or vaginal) probes, surface electrodes for EMG-feedback, rubber or adhesive electrodes for electro stimulation.

GymnaUniphy NV/SA Pasweg 6A | B-3740 Bilzen | +32 (0)89 510 510 | f +32 (0)89 510 511 | info@gymna.com
BTW BE 0434.768.252 TVA | Steuernummer 42/678/1813/8 | UST-ID Nr. DE229596415 | Ondernemingsnr. 0434768252

Algemene verkoopvoorwaarden; zie achterzijde. Conditions de vente: voir verso. General conditions of sale: see back. Allgemeine Verkaufsbedingungen: siehe Rückseite.

Amrite Ganai
Dr. Amrita Ganai
डॉ० अमृता गानै / Dr. Amrita Ganai
अपर-आचार्य / Additional Professor
स्त्री एवं प्रसूति रोग विभाग
Department of Obstetrics & Gynaecology
एमएस ब्रह्मपिकेश / AIIMS, Rishikesh