



अखिल भारतीय आयुर्विज्ञान संस्थान, ऋषिकेश, उत्तराखंड-२४९२०३

All India Institute of Medical Sciences, Rishikesh, Uttarakhand-249203

हृदय-वक्ष एवं वाहिका शल्य चिकित्सा विभाग

Department of CTVS

Date – 10.06.2021

## CORRIGENDUM

To

Tender Office  
AIIMS Rishikesh

**Subject:- Corrigendum for Paediatric Sternal saw Tender.**

***Ref. : - Tender ref no. 24/RT/CTVS/Sternal saw/619 /2021-RISH (ADMN) for Paediatric Sternal saw.***

Dear Sir/Madam,

In view of representation received from Stryker via e-mail, following specifications are added for redo saw hand piece to the original specification.

33. Details of redo saw hand piece are as follows.

- Should have variable speed control on hand piece.
- Should have V-Notch blade mount for cutting blade for better grip and easy maintenance.
- Microprocessor controlled Hand piece Can be calibrated for the consistence performance.
- Saw Noise level should not be more than 100 dB.
- Weight of hand piece with battery should be not more than 5 lbs.
- Should have tool less mounting of cutting blades.
- Should have DC brush less motor for low maintenance.
- Must be not required any lubrication for lifetime for easy maintenance
- Should be autoclavable.
- Should have safe mode for safety on hand piece.
- Should have Pistol grip Hand piece for more comfort of surgeon hand.

Thanking You.

*Anish Gupta*  
10/06/2021  
डा. अनिश गुप्ता /Dr. Anish Gupta  
सहायक आचार्य, /Assistant Professor  
हृदय-वक्ष एवं वाहिका शल्य चिकित्सा विभाग  
Department of CTVS  
एम्स, ऋषिकेश/AIIMS, Rishikesh

*Dr. Anish Gupta*  
10/06/2021  
डा. अनिश गुप्ता  
सह-आचार्य, /Associate Professor  
हृदय-वक्ष एवं वाहिका शल्य चिकित्सा विभाग  
Department of CTVS  
एम्स, ऋषिकेश/AIIMS, Rishikesh

Encl :- 01.

Department of CTVS, 6<sup>th</sup> Floor, Block- A, Medical College Building  
All India Institute of Medical Sciences (AIIMS), Rishikesh – 249203,  
Uttarakhand, India

Website: www.aiimsrishikesh.edu.in  
E- mail: ctvs@aiimsrishikesh.edu.in

CTVS office Phone: +91-0132462514  
Tele-Medicine number: 7417051576 (M)