

CORRIGENDUM-2

Point Sr. No. 48

For: - Endo suturing instrument for endoscopic suturing and knot tying, with one handed operation along with easy needle transfer between jaws. Maximum jaw opening of 19 mm with an option to hold different sutures. Should fit through a 10 mm trocar.

Read As: - Deleted

Point Sr. No. 49

For: - Polysorb 2/0 size, 48-inch-long suture with T shaped needle compatible with endo suturing instrument

Read As :- Deleted

Point Sr. No. 50

For: - Surgidac 2/0 size, 48-inch-long suture with T shaped needle compatible with endo suturing instrument

Read As:- Deleted

Point Sr. No. 64

For: - 6-gram Dropper Bottle Single -use Anti fog solution for providing a fog-free endoscope.

Read As :- Deleted

Point Sr. No. 97

For :- Laparoscopic GIA Linear Cutter roticulating reload with inbuilt knife, 60mm, with Three Different Leg Length for Medium to Thick Tissue range

Read As :- Deleted

Point Sr. No. 98

For :- Laparoscopic GIA Linear Cutter roticulating reload with knife, 60mm, with Three Different Leg Length for Vascular to Medium Tissue range.

Read As:- Deleted

Point Sr. No. 99

For:- Laparoscopic GIA Linear Cutter roticulating reload with knife, 60mm, with Three Different Leg length for Extra Thick Tissue range.

Read As:- Deleted

Point Sr. No. 100

For:- Laparoscopic GIA Linear Cutter roticulating reloads curved tip reload with knife, 60mm, with Three Different Leg Length for Medium to Thick Tissue range

Read As:- Deleted

Point Sr. No. 101

For: - Laparoscopic GIA Linear Cutter roticulating reload with knife, 45mm, with Three Different Leg Length for Vascular to Medium Tissue range.

Read As: - Deleted

Point Sr. No. 102

For: - Laparoscopic GIA Linear Cutter roticulating reload with knife, 45mm, with Three Different Leg Length for Extra Thick Tissue range..

Read As: - Deleted

Point Sr. No. 103

For: - Laparoscopic GIA Linear Cutter roticulating reloads curved tip reload with knife, 45mm, with Three Different Leg Length for Medium to Thick Tissue range

Read As: - Deleted

Point Sr. No. 104

For: - Laparoscopic GIA Linear Cutter roticulating reload with knife, 30mm, with Three Different Leg Length for Vascular to Medium Tissue range.

Read As: - Deleted

Point Sr. No. 105

For: - Laparoscopic GIA Linear Cutter roticulating reloads curved tip reload with knife, 30mm, with Three Different Leg Length for Medium to Thick Tissue range

Read As: - Deleted

Point Sr. No. 106

For: - Laparoscopic Articulating Linear 45 mm purple reload preloaded with tissue reinforcement material for Endo GIA stapler. It has the blade in the reload itself with 3 rows of Staples on both sides with progressive staple heights of 3.0 mm, 3.5 mm and 4.0 mm with a stepped cartridge face

Read As: - Deleted

Point Sr. No. 107

For: - Laparoscopic Articulating Linear 45 mm black reload preloaded with tissue reinforcement material for Endo GIA stapler. It has the blade in the reload itself with 3 rows of Staples on both sides with progressive staple heights of 4.0 mm, 4.5 mm and 5.0 mm with a stepped cartridge face

Read As: - Deleted

Point Sr. No. 108

For: - Laparoscopic Articulating Linear 60 mm purple reload preloaded with tissue reinforcement material for Endo GIA stapler. It has the blade in the reload itself with 3 rows of Staples on both sides with progressive staple heights of 3.0 mm, 3.5 mm and 4.0 mm with a stepped cartridge face

Read As: - Deleted

Point Sr. No. 109

For: - Laparoscopic Articulating Linear 60 mm black reload preloaded with tissue reinforcement material for Endo GIA stapler. It has the blade in the reload itself with 3 rows of Staples on both sides with progressive staple heights of 4.0 mm, 4.5 mm and 5.0 mm with a stepped cartridge face

Read As: - Deleted

Point Sr. No. 110

For: - ENDO linear stapler - 1. It should be a 30mm disposable reloadable laparoscopic linear stapler.

2. It should be reloadable for up to 7 times for a total of 8 applications
3. It should have a non-reflective shaft, with a diameter of 12mm
4. The shaft should be having a 360-degree rotation option
5. The gun handle and closure lever should be with optical indicator and audible feedback, when closed

Read As: - Deleted

Point Sr. No. 111

For: - Endo linear Reloads compatible for ENDO Linear Stabler-1. The size of open staple should be 3 x 2.5mm 2. The loading unit should contain 3 staggered rows of titanium staples with approximate B-shape closure of 1.0mm

Read As: - Deleted

Point Sr. No. 112

For: - Single powered handle compatible with all Endo GIA 30mm, 45mm, 60mm, curved tip and reinforced reloads with unlimited points of articulation between the 45-degree left and right limit. It should come with a rechargeable battery and the product should last for minimum 100 cases or 900 firings whichever is earlier. It should come with reusable adapters, compatible for general laparoscopic surgery. It should be compatible with 30, 45 & 60 MM Endo GIA reloads

Read As: - Deleted