

# All India Institute of Medical Sciences

Veerbhadra Marg, Pashulok,

Rishikesh-249203



## Rate Contract document for procurement of

### General Consumable Items

### AIIMS, Rishikesh

Ref. No.	:	24/RT/124(iii)/2018-Rish(Admn)
Publishing Date	:	24/08/2018 at 9.00 AM
Pre-Bid Meeting	:	18/09/2018 at 3.00 PM
Bid Submission Start Date	:	24/09/2018 at 9.00 AM
Last Date of Bid Submission	:	03/10/2018 at 3.00 PM
Bid Opening	:	04/10/2018 at 3.30 PM

Tender documents may be downloaded from institute's web site [www.aiimsrishikesh.edu.in](http://www.aiimsrishikesh.edu.in) (for reference only) and CPPP site <https://eprocure.gov.in/eprocure/app>

**TENDER NOTICE**

**General Consumable Items**

**AIIMS, Rishikesh, Virbhadra Marg, Rishikesh, Dehradun**

**Dated: 24/08/2018**

1.E-tenders in Two Bids (Technical & Financial) are invited on behalf of the Director, All India Institute of Medical Sciences, Rishikesh from interested and eligible manufacturer or their authorised distributors/dealers, for providing **General Consumable Items on rate contract basis for a period of two years at AIIMS Rishikesh.**

2. Bids shall be submitted online only at CPPP website: <https://eprocure.gov.in/eprocure/app>.

3. The complete bidding process is online. Bidders should be possession of valid digital Signature Certificate (DSC) for online submission of bids. Prior to bidding DSC need to be registered on the website mentioned above. For any assistance for e-bidding process, if required, bidder may contact to the helpdesk at 0135-2462915.

4. Tenderer/Contractor/Bidders are advised to follow the instructions provided in the 'Instructions to the Contractors/Tenderer/Bidders for the e-submission of the bids online through the Central Public Procurement Portal for e Procurement at <https://eprocure.gov.in/eprocure/app>'.

5. Bid documents may be scanned with 100 dpi with black and white option which helps in reducing size of the scanned document.

6. Hard Copy of earnest money deposit and tender fee etc. must be delivered to AIIMS, Rishikesh on or before last date/time of Bid Submission as mentioned above. The bid without EMD and tender fee will be summarily rejected.

7. The tender shall be submitted online in two part, viz., technical bid and financial bid. All the pages of bid being submitted must be signed and sequentially numbered by the bidder irrespective of nature of content of the documents before uploading.

**The offers submitted by Telegram/Fax/email shall not be considered. No correspondence will be entertained in this matter.**

8. Any future clarification and/or corrigendum(s) shall be communicated through <https://eprocure.gov.in/eprocure/app> or through our website: [www.aiimsrishikesh.edu.in](http://www.aiimsrishikesh.edu.in). The bidders are required to regularly check the website to know about any/all such corrigendum(s) as only these bids, taking care of such corrigendum(s) shall be considered for finalisation of the tender. The technical bids shall be opened on the same day at 03.30 PM at Tender Opening Room, AIIMS, Rishikesh. In the event of any of the above mentioned date being declared as a holiday / closed day, the tenders will be opened on the next working day at the appointed time

9. Bidder should necessarily enclose a covering letter mentioning a summary of applied document with proper numbering. Secondly, bid should also attach a compliance sheet as per specification mentioning that they are complying to all specification or have any variation.

10. The pre bid conference would be held on **18/09/2018 at 03.00 PM** in the office of Tender Opening Room, AIIMS, Rishikesh. All firm's representative who are attending the pre bid meeting, shall produce an authorization letter from their firm on the firm's letter head. They are required to put their query in writing before the committee and submit the same in writing on its letter head.

11. All the duly filled/completed pages of the tender should be given serial /page number on each page and signed by the owner of the firm or his Authorized signatory. In case the tenders are signed by the Authorized signatory, a copy of the power of attorney/authorization may be enclosed along with tender. A copy of the terms & conditions shall be signed on each page and submitted with the technical bid as token of acceptance of terms & conditions. Tender with unsigned pages/incomplete/partial/part of tender if submitted will be rejected out rightly.

12. Any omission in filling the columns of Financial Bid form (Schedule of Rates) shall debar a tender from being considered. Rates should be filed up carefully by the tenderer. All corrections in this schedule must be duly attested by full signature of the tenderers. The corrections made by using fluid and overwriting will not be accepted and tender would be rejected.

13.(i) **Bid Security:** -The bidder shall pay the respective amount of Bid Security (EMD) of **Rs. 50,000/- (Rs. Fifty Thousand Only)** along with the Technical Bid by way of demand draft/FD/TD/CD in favour of "AIIMS, Rishikesh" drawn on any Nationalized Bank/ Scheduled Bank and payable at Rishikesh and must be valid for (6) six month. Bids received without tender fees and Earnest Money deposit (EMD) shall stand rejected and thus shall not be considered for evaluation etc. at any stage. The original EMD will be put in cover-I containing with Technical bid.

- a) The Public Sector Undertaking of the Central/State Govt./MSEs/MSME/Registered with Central Purchase Organization are exempted from furnishing Earnest Money along with tender, **subject to submission of its valid proof.**
- b) The firms Registered with DGS & D/SSI and any approved source of Centre/States Govt. are not exempted from furnishing Earnest Money in so far as this institute is concerned.
- c) Earnest Money deposited with AIIMS, Rishikesh in connection with any other tender enquiry even if for same/similar material / Stores by the tenderer will not be considered against this tender.

(ii) EMD is required to protect the purchaser against the risk of Bidders conduct. The EMD will be forfeited if the bidder withdraws or amends its tender or impairs or derogates from the tender in any respect within the period of validity of its tender or if it comes to the notice that the information/ documents furnished in its tender is incorrect or false.

14. The bid security (EMD) without interest shall be returned to the unsuccessful bidders after finalization of contract with successful bidder.

15. Bidders are not allowed to submit more than one bid for the same/similar tendered item else all his bids shall be cancelled thereby making him disqualified in addition to the forfeiture of the EMD.

16. The successful bidders have to execute a contract on Indian non judicial stamp paper of Rs.100/- (Rupees one hundred only) within twenty-one (21) days from the date of award of this tender in his favour and also required to furnish the **Security Deposit of Rs. 2,00,000.00 (Rs. Two Lac Only)** in the form of FD/BG/TD/CD for **three months extra of the contract period** from any Nationalised/Schedule bank duly pledged in favour of AIIMS, Rishikesh & payable at Rishikesh only. The EMD deposited by successful bidder may be adjusted towards Security Deposit as demanded above after its validation for the required period. If the successful bidder fails to furnish the full security deposit or difference amount between Security Deposit and EMD within 21 (twenty-one) days after the issue of **Letter of Award** of Work, his bid security (EMD) shall be forfeited and award of tender in suppliers favour automatically stands terminated at his cost & liability, unless time extension has been granted by AIIMS, Rishikesh.

17. The EMD/PBG shall be forfeited if successful bidder fails to supply the goods/equipment in stipulated time or fails to comply with any of the terms & conditions of the contract or fail to sign the contract.

18. The bid shall be valid and open for acceptance by the competent authority of AIIMS Rishikesh for a period of 180 (one hundred eighty) days from the published date of opening of the tenders and no request for any variation in quoted rates and / withdrawal of tender on any ground by bidders shall be entertained. The unilateral withdraw at any stage will cause forfeiture of EMD in addition to any remedy that the purchaser may have under the law.

19. To assist in the analysis, evaluation and computation of the bids, the Competent Authority, may ask bidders individually for clarification of their bids. The request for Clarification and the response shall be in writing but no change in the price or substance of the bid offered shall be permitted.

20. After evaluation, the work shall be awarded normally to the Agency fulfilling all the conditions of the contract and who has quoted the lowest rate as per financial bid after complying with the all the Acts / provisions stated / referred to for adherence in the tender.
21. The competent authority of AIIMS, Rishikesh reserves all rights to accept or reject any/ all tender(s) without assigning any reason. It can also impose/relax any administrative term and condition/specifications of the tender enquiry after due discussion in pre-bid conference. This will be communicated and shown over the website of the Institute. No representation will be considered after pre-bid meeting and **bidders may ensure its queries only in pre-bid meeting**. AIIMS, Rishikesh also reserves the right to reject any bid which in his opinion is non-responsive or violating any of the conditions/specifications without any liability to any loss whatsoever it may cause to the bidder in the process.
22. Tender must be submitted on the prescribed Tender Form otherwise tender will be cancelled straightway.
23. The tender form is not transferable.
24. Canvassing in any form is strictly prohibited and the tenderers who are found canvassing are liable to have their tenders rejected out rightly.
25. It is required by all concerned, namely the Bidders/Suppliers, as the case may be to observe highest standard of ethics during the procurement and execution of this Tender.
26. Installation at consignee's site should be free of cost immediately on arrival of equipment at consignee's site.
27. In case the quality of goods supplied are not in conformity with the standard given in tender and as per the samples supplied or the supplies are found defective at any stage these goods shall immediately will be taken back by the supplier and will be replaced with the tender quality goods, without any delay. The competent authority reserves all rights to reject the goods if the same are not found in accordance with the required description / specifications and liquidates damages shall be charged in addition to the cost of re- tender. The supplier is required to provide the demonstration of equipment at AIIMS premises to the AIIMS representatives for its evaluation as per the specification & desired functionality standard. However, a submission of videography displaying functionality as per tender specification may also be considered by the AIIMS, representative.
28. It must be mentioned clearly whether bidder is a manufacturer/sole distributor/ sole agent for the items for which he is quoting.
- a. **Manufacturer** must add a certificate that item(s) is manufactured by them as per range of products.
  - b. **Sole Manufacturers** must add a certificate that they are the sole manufacturer of the Item for which they are quoting in this tender enquiry & item is /are their proprietary Item in India. The rate certificate is also required from the sole manufactures that the Rates quoted are the same as they quote to other State/Centre Govt./reputed Private Organisation and DGS&D rate for the similar item(s) and these are not higher than those quoted by them.
  - c. **Authorized agents** must add authority letter from their Manufacturer/Principals on the letter head of the manufacturer/principals signed by a competent person and comes in proforma given in attach must duly supported by a notarised affidavit on Indian Non Judicial Stamp Paper of Rs.10/- (Rupees ten only) that they are quoting Rates on behalf of them. The authorization letter must give/mention the purpose for which it is allowed. The validity period of the authorization letter must be mentioned in the authority letter otherwise tender will be liable to rejection.
29. In case the bidder on whom the supply order has been placed, fails to make supplies within the delivery schedule and the purchaser has to resort risk purchase, the purchaser (AIIMS, Rishikesh) may recover from the tender the difference between the cost calculated on the basis of risk purchase price and that calculated

on the basis of rates quoted by tenderer. In case of repeated failure in supplying the order goods the supply order may be cancelled and bid security deposit will be forfeited.

30. The bidders should have furnished a copy of GST/S.T. /C.S.T./VAT registration number, the State / U.T. of registration and the date of such registration. Tenders not complying with this condition will be rejected.

31. **Turnover provisions:** - (i) The tenderers should submit along with the tender, a photo state copy of the last three years Annual Accounts with Audit certificate by Chartered Accountant, Income Tax returns and a copy of current valid income tax clearance certificate (IT CC), otherwise bidder will not be considered for administrative evaluation (in evaluation of Technical bid) and will be declared **disqualified** in technical evaluation.

(ii) In case of bidder falls under Section 44AD/44ADA/44AE of Income Tax shall be required to submit turnover certificate verified by Chartered Accountant with his Registration number issued by Institute of Chartered Accountants of India.

(iii) In case of supplier is an Indian Agent, the firm can submit copies of purchase orders issued in favour of firm (*As the payment is made through LC directly to foreign manufacturer and equipment payment does not exist in the book of account of the supplier*) in support to its turnover whatever amount is getting short.

(iv) There will be relaxation on turnover on *case to case* basis for **Start-up firms** registered by Government of India under Start-ups scheme as per orders of Ministry of Commerce, Government of India.

32. **Sample/demonstration:** - In case, the item required prior submission of sample/ performing demonstration, tenderer will have to submit sample/perform demonstration of the equipment/item to the competent authority of the institute, the bidder will have to born all the expenses for the same. Non submission of sample/non performing demonstration will **disqualify** the bidder in the technical bidding process and financial bid of the bidder will not be opened.

33. The tenderer hereby guarantees that the equipment supplied to the Institute (purchaser) under the Contract shall be of the best quality/latest version and workmanship and new in all respects and shall be strictly in accordance with the specification and particulars contained/ mentioned in the Tender Document. The date of manufacturing of the equipment/goods supplied will not be more than 3 (Three months) old. The tenderer will have further guarantees that the said equipment would continue to conform to the description and quality aforesaid for a period of five (5) years guarantee period (As per MoH&FW guidelines), from the date of installation of the said equipment to the purchaser and notwithstanding the fact that the Purchaser (Inspector) may have inspected and /or approved the said equipment, if during the aforesaid period of five years the said equipment be discovered not to conform to the description and quality as required as per specification or not giving satisfactory performance or have deteriorated, the decision of the Purchaser in that behalf shall be final and binding on the tenderer and the Purchaser shall be entitled to call upon the tenderer to rectify the equipment or such portion thereof as is found to be defective by the purchaser within a reasonable period or such specified period as may be allowed by the purchaser in his discretion on/an application made thereof by the tenderer and in such an event, the above mentioned warranty period shall apply to the equipment replaced from the date of replacement thereof. In case of failure of the tenderer to rectify or replace the equipment, within specified time, the purchaser shall be entitled to recover the cost with all expenses from the tenderer for such defective equipment.

34. Full description & specifications, make/brand and name of the manufacturing firm must be clearly mentioned in the tender, failing which, the tender will not be considered. The tenderer must also mention whether the goods are imported / indigenous. Descriptive literature / catalogues must be attached with the tender in original, failing which, tender may be disqualified.

**35. Force Majeure:** Any failure or omission to carryout of the provisions of this supply by the supplier shall not give rightfor any claim by supplier and purchaser to one against the other, if such failure or omission arise from an act of God which shall include all acts of natural calamities from civil strikes compliance with any statistics and or requisitions of the Government lockout and Strikes, riots, embargoes or from any political or other reasons beyond the suppliers control including war (whether declared or not) civil war or state of incarceration provided that notice of the occurrence of any event by either party to the other shall be within two weeks from the date of occurrence of such an event which could be attributed to force majeure. Any delay due to Force Majeure will not be attributable to the either of the parties.

36. The equipment installed should be up for 95% of the total warranty time. If the equipment is down for more than 5% suitable action shall be taken against the supplier including imposition of penalty as deemed fit.

37. If there is a close system the tenderer shall ensure and will have to submit an affidavit on Indian Non Judicial stamp paper of Rs.10/- along with technical bid that spare parts and consumables for these equipment's/instruments/item will be available at reasonable fixed rates for next 10 (ten) years, such rates should not be more than the rates supplied to institutes of national importance.

38. The successful Bidder shall at all times agree to indemnify and keep indemnified the purchaser against all losses, damages which may arise in respect of action/inactions of such Bidder or breach of any term of this tender by such Bidder. All claims regarding indemnity shall survive the termination of the contract with such Bidder.

39. In case the vendor fails to supply the spare parts or fails to provide the agreed maintenance during the prescribed period, as per the terms of contract, the purchaser is automatically entitled to procure the required parts and hire services from the market at the risk and cost of the vendor, such inability of bidder will entail forfeiture the security deposit. The purchaser also reserves the right to terminate the contract on immediate notice, if the vendor fails to comply with this clause for more than one instance.

**40. Liquidated damage/demerge:** - The time for the date of delivery/ dispatch stipulated in supply order shall be deemed to be the essence of the contract and if the supplier fails to deliver or dispatch any consignment within the period prescribed for such delivery or dispatch in the supply order, liquidated damages may be deducted from the bill @ 0.5% per week subject to maximum of 10% of the value of the delayed goods or services under the contract. The competent authority of the institute may also cancel the supply at the cost & liability of the supplier. In such a case, bid security of the supplier shall stand forfeited. The supply of equipment must be in single consignment, inclusive of all parts & accessories in adherence to the specification so as to make the equipment fully functional at the time of the installation. No installation repeat shall be signed in case of absence of any part as per the specification.

**41. Legal Jurisdiction:** -The Courts at Rishikesh/ Dehradun alone and no other Court will have the jurisdiction to try the matter, dispute or reference between the parties arising out of this tender/supply Order/contract.

**Applicable Law:**

42. The contract shall be governed by the laws and procedures established by Govt. of India, within the framework of applicable legislation and enactment made from time to time concerning such Commercial dealings / processing.

43. Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Rishikesh/Dehradun, Uttarakhand India only.

44. Except as otherwise provided under this Contract for immediate termination of the Contract, in the event of a disputes which may be arising out of the execution of the tender contract, the matter will be referred to the Deputy Director (Administration). Appeal against the decision of the DDA will lie to the Director, AIIMS Rishikesh and his decision shall be final and binding upon both the parties.

I / We hereby accept the terms and Conditions given in the tender

---

(Signature & Stamp of the bidder)

*Note- Please sign each page of document including terms & conditions & tender*

**(B) Financial terms and conditions**

1. Rates are strictly required to be offered/quoted on the prescribed **“Financial Bid format”**. Financial bid submitted without prescribed format may not be considered and will be deemed improper subject to the condition that there is a requirement to mention other item which are not mentioned in the said format.
2. Rates quoted should be inclusive of all applicable taxes, packing, forwarding, postage and transportation charges at for AIIMS Rishikesh (Site of installation/Use).
3. The supplier has to submit a notarised affidavit on Indian Non Judicial Stamp Paper of Rs.10/- that the bidder has not quoted the price higher than previously supplied to any government Institute/Organisation/reputed Private Organisation or DGS&D rate in recent past. Therefore, if at any stage it has been found that the supplier has quoted lower rates than those quoted in this tender; the Institute (the purchaser) would be given the benefit of lower rates by the Supplier and any excess payment if any, will become immediately payable to the AIIMS, Rishikesh. If such affidavit is not submitted, tender will be out rightly rejected.**(Part of technical bid)**
4. If the price of the contracted articles is/ are controlled by the Government, in no circumstances the payment will be higher than the controlled rate.
5. Tender will be regarded as constituting an offer open to acceptance in whole or in part at the discretion of the competent authority of the institute for a period of 180 days (6 months) valid from the date of opening of the tender by the committee.
6. Order shall be issued for tentative annual requirement on actual need basis. Bills in triplicate for the items supplied by the selected firm(s), should be raised for payment. Payment shall be released after it is ensured that the items/quantity and quality of items supplied are to the entire satisfaction of this office and accepted. If any item is found to be defective, or not of the desired quality, the same shall be replaced immediately, for which no extra payment shall be made by AIIMS, Rishikesh.
7. The corrections made by using fluid and overwriting will not be accepted and tender would be rejected.
8. **Tender Currencies:** The bidders are required to quote in Indian Rupees only.
9. **Sales Tax:** - If a bidder asks for sales tax/GST/CST / VAT/CENVAT, Service Tax and Works Contract Tax to be paid extra, the rate and nature of sales tax applicable should be shown separately. The GST/CST / VAT/CENVAT, Service Tax and Works Contract Tax will be paid as per the rate at which it is liable to be assessed or has actually been assessed provided the transaction of sale is legally liable to sales tax/ GST/CST / VAT/CENVAT, Service Tax and Works Contract Tax and is payable as per the terms of the contract.
10. **Octroi Duty and Local Duties & Taxes:** - Normally, goods to be supplied to Government departments against Government contracts are exempted from levy of town duty, Octroi duty, terminal tax and other levies of local bodies. However, on some occasions, the local bodies (like town body, municipal body etc.) as per their regulations allow such exemptions only on production of certificate to this effect from the concerned Government department. Keeping this in view, the supplier shall ensure that the goods to be supplied by the supplier against the contract placed by the AIIMS, Rishikesh are exempted from levy of any such duty or tax and, wherever necessary, obtain the exemption certificate from the AIIMS, Rishikesh. However, if a local body still insists upon payment of such local duties and taxes, the same should be paid by the supplier to the local body to avoid delay in supplies and possible demurrage charges and obtain a receipt for the same. The supplier should forward the receipt obtained for such payment to the AIIMS, Rishikesh to enable the AIIMS, Rishikesh reimburse the supplier and take other necessary action in the matter.
11. **Payment terms:** - The payment shall be made to the supplier on the satisfactory receipt of the goods ordered. Advance payment shall not be made under any circumstances. The payment shall be made to the suppliers A/C through NEFT/RTGS mode as per the details provided by the supplier.

12. **L1 Clause: Rate contract will be awarded to firm who will quote lowest per unit rates for items as per list provided in .pdf format separately.**

NO DEMURRAGE / WHARFAGE CHARGES WILL BE PAYABLE BY THE INSTITUTE UNDER ANY CIRCUMSTANCES. NO ADVANCE PAYMENT WILL BE PAYABLE FOR CUSTOM CLEARANCE/ FREIGHT/INSURANCE ETC

**Note: In case of any dispute regarding award of tender, decision of AIIMS Administration would be final.**

I / We hereby accept the terms and Conditions given in the tender

-----  
(Signature & Stamp of the bidder)

*Note- Please sign each page of document including terms & conditions & tender*



**“General Consumable Items”**

**For**

**AIIMS, Rishikesh**

**TECHNICAL BID**

**(Documents to be attached in the “Technical Bid”)**

1. Name & Address of the manufacturer and their authorised dealers/ distributors/Agency with phone number, email, name and telephone/mobile	
2. Specify your firm/company is a manufacturer/ authorised dealer/ distributor/ Agency	
3. Whether the signature on each page has been made by the bidder or not.	
4. Name, Address & designation of the authorized person (Sole proprietor/partner /Director)	
5. Have you previously supplied these items to any government/ reputed private organization? If yes, attach the relevant poof.	
6. Please provide a notarised affidavit on Indian Non Judicial stamp paper of Rs. 10/- that you have not quoted the price higher than previously supplied to any government Institute/Organisation/reputed Private Organisation or DGS&D rate in last one year. If you don't fulfil this criteria, your tender will be out rightly rejected.	
7. Please attach copy of last three years' of Income Tax Return	
<b>8. Turnover</b> a. Please attach balance sheet (duly certified by Chartered Accountant) for last three (3) years (Attach copy of annual minimum turnover which should not be less than 50 lakh duly certified by the Chartered Accountant) b. The bidder falls under Section 44AD/44ADA/44AE of Income Tax Act shall be required to submit turnover certificate verified by Chartered Accountant with his Registration number issued by Institute of Chartered Accountants of India. c. Indian Agent can submit its copy of POs of LC cases, in support of its amount getting short in required turnover. d. Start-ups may submit its Start-up Registration for consideration ( <i>Relaxation in turnover can be considered as the case may be, subject to fulfilment of other conditions. However, it will not mandatory</i> )	
9. PAN No. (Please attach copy)	
10. GST/VAT/Service Tax Registration Number. (Please attach copy)	
11. Acceptance of terms & conditions attached (Yes/No). Please sign each page of terms and conditions as token of acceptance and submit as part of tender document with technical bid. Otherwise your tender will be rejected.	
12. Power of Attorney/authorization for signing the bid documents (Not required in case of sole-proprietorship.)	
13. Please submit a notarised affidavit on Indian Non judicial stamp paper of Rs. 10/- that no case is pending with the police against the Proprietor/firm/partner or the Company (Agency). Indicate any convictions in the past against the Company/firm/partner.	
14. Please declare that proprietor/firm/company has never been black listed/debarred by any organization. An oath certificate to this effect may be enclosed on Rs.10 notarised stamp paper.	
15. Please submit two performance certificate from your two different customers to whom you have supplied such type of consumables in previous 3 years	
16. Details of the FD/DD/TD/CD of bid security (EMD)	Detail of cost of Tender for Rs. 1180/- (if downloaded from website)
FD/DD/TD/CD No:	<b>DD No.</b>
Date:	<b>Date:</b>
Payable at-	<b>Payable at-</b>

**Undertaking**

1. I/We have read and understood the contents of the Tender and agree to abide by the terms and conditions of this Tender.
2. I/We shall supply the items of requisite quality and quantity at given rate in timely manner.
3. I/We also confirm that in the event of my/our tender being accepted, I/we hereby undertake to furnish within 15 days, Bank Guarantee/ Performance Security after the issue of Purchase Order, as applicable, in the format to be provided by AIIMS Rishikesh in addition to execution of a Contract as pre-condition for obtaining the supply orders.
4. I/We further undertake that none of the Proprietor/Partners/Directors of the firm was or is Proprietor or Partner or Director of any firm with whom the Government have banned /suspended business dealings. I/We further undertake to report to the AIIMS Rishikesh immediately after we are informed but in any case not later 15 days, if any firm in which Proprietor/Partners/Directors are Proprietor or Partner or Director of such a firm which is banned/suspended in future during the currency of the Contract with you.
5. I/We undertake that the information given in this tender are true and correct in all respect and I/We hold the responsibility for the same.

(Signature of the Bidder)

Name:

Designation with Seal of the Firm:

AIIMS Rishikesh tender Enquiry No. F.No 24/RT/124(iii)//2018-RISH (ADMN)  
**MANUFACTURER's / PRINCIPAL's AUTHORIZATION FORM**

(Clause 11 (c) of **other terms and conditions** of the tender)

To

The Administrative Officer,  
All India Institute of Medical Sciences  
Rishikesh

Dear Sir,

TENDER: \_\_\_\_\_.

we, \_\_\_\_\_, who are established and reputable manufacturers of \_\_\_\_\_, having factories at \_\_\_\_\_ and \_\_\_\_\_, hereby authorize Messrs. \_\_\_\_\_ (*name and address of agents*) to bid, negotiate and conclude the contract with you against Tender No. \_\_\_\_\_ for the above goods manufactured by us.

We hereby extend our full guarantee and warranty as per the conditions of tender contract for the goods offered for supply against this tender by the above firm.

The authorization is valid up to \_\_\_\_\_

Yours faithfully,

(Name)

For and on behalf of Messrs. \_\_\_\_\_  
(*Name of manufacturers*)/Principal.

AIIMS Rishikesh tender Enquiry No. F.No 24/RT/124(iii)//2018-RISH (ADMN)  
**BANK GUARANTEE FORM FOR PERFORMANCE SECURITY/ CMC**  
**SECURITY**

To

**The Administrative Officer**

All India Institute of Medical Sciences  
Rishikesh,  
Virbhadra Marg,  
Rishikesh-249201

WHEREAS \_\_\_\_\_ (Name and address of the supplier) (Hereinafter called “the supplier”) has undertaken, in pursuance of contract no \_\_\_\_\_ dated \_\_\_\_\_ to supply (description of goods and services) (herein after called “the contract”). AND WHEREAS it has been stipulated by you in the said contract that the supplier shall furnish you with a bank guarantee by a scheduled commercial bank recognised by you for the sum specified therein as security for compliance with its obligations in accordance with the contract; AND WHEREAS we have agreed to give the supplier such a irrevocable bank guarantee;

NOW THEREFORE we hereby affirm that we are guarantors and responsible to you unconditionally, on behalf of the supplier, up to a total of. \_\_\_\_\_ (Amount of the guarantee in words and figures), and we undertake to pay you, upon your first written demand declaring the supplier to be in default under the contract and without cavil or argument, any sum or sums within the limits of (amount of guarantee) as aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

We hereby waive the necessity of you to first demanding the said amount of guarantee from the supplier before raising the demand with us. You may directly raise the demand with us, without asking the supplier for the same.

We further agree that no change or addition to or other modification of the terms of the contract to be performed there under or of any of the contract documents which may be made between you and the supplier shall in any way release us from any liability under this guarantee and we hereby waive notice of any such change, addition or modification.

This guarantee will not be changed due to change in the constitution of the bank or the supplier.

This guarantee shall be valid up to 65 months from the date of satisfactory installation of the equipment i.e. up to --  
----- (indicate date).

.....  
(Signature with date of the authorized officer of the Bank)

.....  
Name and designation of the officer

.....  
Seal, name & address of the Bank and address of the Branch

**Instructions for Online Bid Submission:**

The bidders are required to submit soft copies of their bids electronically on the CPP Portal, using valid Digital Signature Certificates. The instructions given below are meant to assist the bidders in registering on the CPP Portal, prepare their bids in accordance with the requirements and submitting their bids online on the CPP Portal.

More information useful for submitting online bids on the CPP Portal may be obtained at: <https://eprocure.gov.in/eprocure/app>.

### **REGISTRATION**

- Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal (URL: <https://eprocure.gov.in/eprocure/app>) by clicking on the link “Online bidder Enrollment” on the CPP Portal which is free of charge.
- As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.
- Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.
- Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate issued by any Certifying Authority recognized by CCA India (e.g. Sify / nCode / eMudhra etc.), with their profile.
- Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSC’s to others which may lead to misuse.
- Bidder then logs in to the site through the secured log-in by entering their user ID / password and the password of the DSC / e-Token.

### **SEARCHING FOR TENDER DOCUMENTS**

- There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, Organization Name, Location, Date, Value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as Organization Name, Form of Contract, Location, Date, Other keywords etc. to search for a tender published on the CPP Portal.
- Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective ‘My Tenders’ folder. This would enable the CPP Portal to intimate the bidders through SMS / email in case there is any corrigendum issued to the tender document.
- The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.

### **PREPARATION OF BIDS**

- Bidder should take into account any corrigendum published on the tender document before submitting their bids.
- Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid.

- Please note the Number of covers in which the bid documents have to be submitted, the number of documents - including the names and content of each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid.
- Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF / XLS / RAR / DWF/JPG Tender No. : ..... formats. Bid documents may be scanned with 100 dpi with black and white option which helps in reducing size of the scanned document.
- To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use “My Space” or “Other Important Documents” area available to them to upload such documents. These documents may be directly submitted from the “My Space” area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process.

### **SUBMISSION OF BIDS**

- Bidder should log into the site well in advance for bid submission so that they can upload the bid in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.
- The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document.
- Bidder has to select the payment option as “offline” to pay the tender fee / EMD as applicable and enter details of the instrument.
- Bidder should prepare the EMD as per the instructions specified in the tender document. The original should be posted/couriered/given in person to the concerned official, latest by the last date of bid submission or as specified in the tender documents. The details of the DD/any other accepted instrument, physically sent, should tally with the details available in the scanned copy and the data entered during bid submission time. Otherwise the uploaded bid will be rejected.
- Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. If the price bid has been given as a standard BoQ format with the tender document, then the same is to be downloaded and to be filled by all the bidders. Bidders are required to download the BoQ file, open it and complete the white coloured (unprotected) cells with their respective financial quotes and other details (such as name of the bidder). No other cells should be changed. Once the details have been completed, the bidder should save it and submit it online, without changing the filename. If the BoQ file is found to be modified by the bidder, the bid will be rejected.
- The server time (which is displayed on the bidders’ dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.
- The documents being submitted by the bidders would be encrypted using PKI encryption all techniques to ensure the secrecy of the data. The data entered cannot be viewed by unauthorized persons until the time of bid opening. The confidentiality of the bids is maintained using the secured Socket Layer 128 bit encryption technology. Data storage encryption of sensitive fields is done. Any bid document that is uploaded to the server is subjected to symmetric encryption using a system generated symmetric key
- Further this key is subjected to asymmetric encryption using buyers/bid opener’s public keys. Overall, the uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- The uploaded tender documents become readable only after the tender opening by the authorized bid openers.

- Upon the successful and timely submission of bids (i.e. after Clicking “Freeze Bid Submission” in the portal), the portal will give a successful bid submission message & a bid Tender No. : ..... summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.
- The bid summary has to be printed and kept as an acknowledgement of the submission of the bid. This acknowledgement may be used as an entry pass for any bid opening meetings.

**ASSISTANCE TO BIDDERS**

- Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.
- Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk number 0120-4200462, 0120- 4001002.

**List of General Consumable Items**

Sl. No.	NAME OF THE ITEM WITH SPECIFICATION
1	Vibratory Pep Therapy System For Lung Secretion Clearance.-Vibration And Positive Expiratory Pressure Generating Devices For Secretion Clearance Must Have 22Mmod Connection To Accommodate Face Mask Or A Mouth Piece, Expiratory Resistance Dial.
2	Percutaneous Dilatation Tracheostomy Set With Single Stage Dilator-Should Have Scalpel With Number 15 Blade, An Introducer Needle,10ML Syringe For Aspiration,Pre-Dilator,Should Have A Well Marked Extra Length Guiding Catheter,Should Have A Hydrophilic Graduated Single Stage Dilator,Should Have A Cuffed Tracheostomy Tube ,Should Have A Tracheostomy Tube Introducer,Should Have A Flexible Inner Cannula, Should Be Fda/Ece Approval.Sizes 7,8,9Mm
3	Volumetric Incentive Spirometer-In Steps Of 250 ML Calibration And Maximum Tidal Volume Is 2500 ML,Low Flow Siphon With Rattle Indicator,One Way Valve For Inhalation Only,Flexible Pebble Tube With Mouth Piece,Inbuilt Oxygen Port,Should Have Usfda Approval And Ece Certification.
4	Percutaneous Dilatation Tracheostomy Set With Single Stage Dilator-Should Have Scalpel With Number 15 Blade, An Introducer Needle,10ML Syringe For Aspiration,Pre-Dilator,Should Have A Well Marked Extra Length Guiding Catheter,Should Have A Hydrophilic Graduated Single Stage Dilator,Should Have A Cuffed Tracheostomy Tube ,Should Have A Tracheostomy Tube Introducer,Should Have A Flexible Inner Cannula, Should Be Fda/Ece Approval.Sizes 7,8,9Mm
5	Volumetric Incentive Spirometer-In Steps Of 250 ML Calibration And Maximum Tidal Volume Is 2500 ML,Low Flow Siphon With Rattle Indicator,One Way Valve For Inhalation Only,Flexible Pebble Tube With Mouth Piece,Inbuilt Oxygen Port,Should Have Usfda Approval And Ece Certification.
6	Epidural Kit With Catheter Fixation Device -Should Have Tuohy Needles16G/18G,Lor Syringe Should Have An Epidural Catheter With A Smooth Tip, Should Have Catheter Fixation Device,Should Have A Catheter Connector With A Flat Profile For Increased Patient Comfort,Should Have A Flat Filter With A 0.2µm Hydrophilic Supported Membrane That Allows Two Way Filtration. Ece/Usfda Approved
7	Multi Rate Elastomeric Pump - Single-Use, Portable Infusion Pump With Flow Rate Accuracy Of ±10%,Air-Vent Blue End Cap With 0.2Mm Hydrophobic Membrane Forces Air Bubbles Out, While Preventing Medication From Flowing Out, Micro Iv Filter That Has 1.2Mm Hydrophilic - 0.2Mm Hydrophobic Membranes That Expel The Particles And Air From The Infusion Line. Available In Nominal Volume Of 100ML, 275ML,Available With 2 Flow Rates Of 1/2/3/4/5/6/7 ML/Hr And 2/4/6/8/10/12/14 ML/Hr ,Ce Approved
8	Single Rate Elastomeric Pump-Single-Use, Portable Infusion Pump With Flow Rate Accuracy Of ±10%,Air-Vent Blue End Cap With 0.2Mm Hydrophobic Membrane Forces Air Bubbles Out, While Preventing Medication From Flowing Out, Micro Iv Filter That Has 1.2Mm Hydrophilic - 0.2Mm Hydrophobic Membranes That Expel The Particles And Air From The Infusion Line. Available In Nominal Volume Of 100ML, 275ML,Available With Continuous Flow Rate Of 2/ 5/ 8/ 10ML/H, Ce Approved
9	Single Rate Elastomeric Pump With Pca-Single-Use, Portable Infusion Pump With Flow Rate Accuracy Of ±10%,Air-Vent Blue End Cap With 0.2Mm Hydrophobic Membrane Forces Air Bubbles Out, While Preventing Medication From Flowing Out, Micro Iv Filter That Has 1.2Mm Hydrophilic - 0.2Mm Hydrophobic Membranes That Expel The Particles And Air From The Infusion Line. Available In Nominal Volume Of 100ML, 275ML,Available With Continuous Flow Rate Of 2/ 5 ML/H With 2MI Bolus 15Min Lockout Time Ce Approved
10	Ivory North Nasal Endotracheal Tube-High Qualityivorypvc to minimize Risk Oftraumato Nasal Passages,Smoothtip, Murphy Eyeand A Radio-Opaque Blue-Line,Kink-Resistance To Minimize Possibility of Tube Occlusion,Low Pressure-Highvolumesmoothtapered Cuffmadefromvelvetsoft Pvcmaterial, For Minimal Risk Oftrauma,Ithasaninflatablepilotballoontohelpindicatecuffpressure,Havingclearmarkings Suchascuffresting Diameter,Withaspring-Loadedone-Wayvalve,Andablueinflation Line,The Tube Has An Intubation Depth Marker And Clear Depth Markings.Internal And External Diameters Of Tube Are Clearlymarked On The Tubeshould Have Usfda& Implantations Tested ,Available In Sizes: 5,5.5, 6,6.6, 7, 7.5, 8, 8.5, 9



11	Uncuffed Pediatric Tracheostomy Tube -Should Be Made Of Thermosensitive Pvc,Should Have A Soft, Semi Directional Flexible Flange, Should Have Usfda Approved Certificate .Should Be Uncuffed,Sizes (Mm)-3,3.5,4,4.5,5
12	Cuffed Tracheostomy Tube With Sub Glottic Secretion Lumen -Should Have Inbuilt Suction Port Opening Into A Suction Lumen For Suctioning Above The Cuff, Should Have Two Inner Cannulaeshould Have Usfda Approved Certificate-Sizes 6.0Mm,7.0Mm,7.5Mm,8.0Mm,8.5Mm,9.0Mm,10.0Mm.
13	Epidural Set 19G -Should Have Tuohy Needles19G,Lor Syringe Should Have An Epidural Catheter With A Smooth Tip, Should Have A Catheter Connector With A Flat Profile For Increased Patient Comfort,Should Have A Flat Filter With A 0.2µm Hydrophilic Supported Membrane That Allows Two Way Filtration.
14	Reinforced Endotracheal Tubewithoutcuff-Made Of Thermo-Sensitive Clear Pvc With A Smooth Tip And Murphyeye,Spiral Reinforcement Minimizes Crushingor Kinking,Internal And External Diameters Of Tube Are Clearlymarked On The Tube.The Tube Has An Intubation Depth Marker And Cleardepth Markings,Ablue 15Mmconnector.Biocompatible-Tested Inaccordance With Iso10993-1 , Should Have Usfda& Implantations Tested Certificatesavailable In Sizes: 3,3.5, 4,4.5, 5Mm
15	Siliconized Endotracheal Tubewithoutcuff-Made Of Thermo-Sensitive Clear Pvc With A Smooth Tip And Murphyeye,Internal And External Diameters Of Tube Are Clearlymarked On The Tube.The Tube Has An Intubation Depth Marker And Cleardepth Markings, Ablue 15Mmconnector.Biocompatible-Tested Inaccordance With Iso10993-1 Should Have Usfda& Implantations Tested Certificatesavailable In Sizes: 2,2.5,3,3.5, 4,4.5, 5Mm
16	Heat & Moisture Exchangefilters (Hmef)- Hepabacterialandviralfilterwithhmepropertiesofferingoptimumhydrophobicfiltration Performance,Gassampling Port, Forend Tidal Co2Monitoring.Recommendedtidal Volume Range 150- 1200MI,Humidityoutput (Mg H2O Per L): 25.6Per250/18.6Per 50, Even-Pleatedfiltermediumtoallowmaximumfiltrationefficiencywithminimalresistanceto Flow,Lowdeadspace, Equal To 45MI,Singleuse,Sterile Pack
17	Percutaneous Tracheostomy Griggs Forceps,Should Have Scalpel With Number 10 Blade,Should Have An Introducer Needle,Should Have Curved Metallic Forceps For The Blunt Dissection Of Pre Tracheal Tissues,Should Have A 10MI Syringe For Aspiration
18	Should Have 14Fr Pre-Dilator,Should Have Extra-Long Guide Wire,Should Have A Cuffed Tracheostomy Tube With Obturator,Should Have A Tracheostomy Tube Introducer,Should Have Swabs For Cleaning,Should Have A Tracheostomy Tube Holder, Should Have Usfda Approval.Sizes-7,8,9Mm.
19	Abg Heparin Lithium Based Prefilled Syringes With Filter Pro Cap To Expels Out Air Bubble From Blood Sample Available In 1MI
20	Abg Heparin Lithium Based Prefilled Syringes With Filter Pro Cap To Expels Out Air Bubble From Blood Sample Available In 1MI
21	Cuffed Tracheostomy Tube Kit Should Be Made Of Thermosensitive Pvc, High Volume - Low Pressure, Velvet Soft Cuff,Should Have A Hollow Obturator With Rounded Tip For Insertion Over A Guide Wire , Should Have 2 Plain Inner Cannulae ,Cleaning Brush & Tracheostomy Tube Holder, Should Have Usfda/Ece Approved Certificate .Sizes (Mm)- 6,7,7.5,8,8.5,9,10
22	Uncuffedtracheostomy Tube Kitshould Have A Uncuffed Tracheostomy Tube That Is Made Of Thermosensitive Pvc , Tracheostomy Tube Should Have Multiple Fenestrations To Minimize Risk Of Occlusion Of Fenestrations, Should Have 1Plain Inner Cannulae,1 Fenestrated Inner Cannulae ,Cleaning Brush & Tracheostomy Tube Holder, Should Have Usfda/Ece Approved Certificate .Sizes (Mm)-6,7,7.5,8,8.5,9,10

23	Cuffed Fenestrated Tracheostomy Tube Kit-Should Have A Cuffed Tracheostomy Tube That Is Made Of Thermosensitive Pvc , Tracheostomy Tube Should Have Multiple Fenestrations To Minimize Risk Of Occlusion Of Fenestrations, Should Have 1Plain Inner Cannulae,1 Fenestrated Inner Cannulae ,Cleaning Brush & Tracheostomy Tube Holder, Should Have Usfda/Ece Approved Certificate .Sizes (Mm)-6,7,7.5,8,8.5,9,10
24	Silicone Tracheostomy Tube With Aire Cuff-Should Be Made Of Silicone Material, Should Have A Low Pressure, Inflatable, Air Filled Silicone Cuff, Should Be Mri Compatible, Obturator Should Be Guide Wire Compatible.Sizes-2.5-5Mm Pediatric & Adult-5,6,7,8,9Mm.Ece/Usfda Approved
25	Silicone Tracheostomy Tube With Aircuff With Adjustable Flange-Should Be Made Up Of Pure Silicone And Should Be Wire Reinforced, Should Have Extra Length Than Normal Tracheostomy Tube And The Flange Should Be Adjustable As Per The Requirement. Sizes Should Be 6,7,8,9. Ece/Usfda Approved
26	Tracheostomy Hme Filter With Integrated Oxygen Port, Should Have Opening Port Allows Quick And Easy Access For A Suction Catheter To Aid Secreton Removal. Should Be Ece
27	Linear Cutter 55mm having cut line 53 mm and staple line 57 mm with CAM mechanism, intermediate locking position for repositioning of tissue, preloaded with Blue cartridge.
28	LINEAR CUTTER CARTRIDGE 55mm compatible with linear cutter 55mm – Green cartridge having closed staple height of 2.0mm with tissue retaining button.
29	Linear stapler 60mm with controlled tissue compression, parallel jaw closure, longer staple legs, locking trigger in handle & remote retaining pin, heavy wire option.
30	Proximate linear stapler standard cartridge 60mm with yellow (Heavy) reloads compatible with Linear Stapler 60mm.
31	Linear stapler 90mm with controlled tissue compression, parallel jaw closure, longer staple legs, locking trigger in handle & remote retaining pin, heavy wire option.
32	Proximate linear stapler standard cartridge 90mm with yellow (Heavy) reloads compatible with Linear Stapler 90mm.
33	Proximate linear stapler gun 30mm with controlled tissue compression, parallel jaw closure, longer staple legs, locking trigger in handle & remote retaining pin vascular
34	Proximate linear stapler cartridge 30mm Vascular
37	Curved cutter stapler reload cartridge –Green with closed staple height 2 mm.
38	Hemorrhoidal stapler kit consists of 33mm Hemorrhoidal Circular stapler (with fixed anvil, adjustable closed staple height from 0.75 mm – 1.5 mm, staple open leg length of 5.5 mm), Suture Threader, Circular Anal Dilator, Purse-String Suture Anoscope, Suture for Purse String.
39	Proximate Intra-Luminal Staplers with Curved shaft anvil detachable 21mm with Lumen Size12.4mm diameter head, with adjustable staple height (1.0-2.5 mm) for controlled tissue compression, longer staple leg 5.5mm. Non slip grip surface.
40	Proximate Intra-Luminal Staplers with Curved shaft anvil detachable 25mm with Lumen size 16.4mm diameter head, with adjustable staple height (1.0-2.5 mm) for controlled tissue compression, longer staple leg 5.5mm.Non slip grip surface.
41	Proximate Intra-Luminal Staplers with Curved shaft anvil detachable 29mm with lumen size 20.4 diameter head, with adjustable staple height (1.0-2.5 mm) for controlled tissue compression, longer staple leg 5.5mm. Non slip grip surface.
42	Proximate Intra-Luminal Staplers with Curved shaft anvil detachable 33mm with lumen size 24.4 diameter head, with adjustable staple height (1.0-2.5 mm) for controlled tissue compression, longer staple leg 5.5mm. Non slip grip surface.

43	ENDOPOUCH RETRIEVER SPEC BAG-Endopouch specimen bag 4"X6" with retriever handle.
44	Endo-loop for Appendectomy-Endoloop Catgut Chromic
45	Multiple clip applier 24cm with 20 preloaded titanium clips of small (LT100) size.
46	Multiple clip applier 29cm with 30 preloaded titanium clips of medium (LT200) size.
47	Multiple clip applier 33.7 cm with 20 preloaded titanium clips of Large (LT400) size.
48	liga clip applier for small clips with color-coded handle and lateral/transverse groover corresponding to the clip.
49	liga clip applier for medium clips with color-coded handle and lateral/transverse groovercorresponding to the clip.
50	Endoscopic Linear Cutter Cartridge having 45mm staple line, 42 mm cut line,Green cartridge having closed staple height of 2.0mm, compatible with Endoscopic articulating linear cutter 45mm.
51	Endoscopic linear cutter having 60mm staple line, 57 mm cut line, with single handed natural articulation (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 4- stroke firing with stroke indicator, 360 degree rotation.340 mm shaft length – (Normal Length).
52	Battery Operated 60mm Articulating Endo Cutter with a disposable battery pack, for enhanced distal tip stability while firing, 360 degree rotation shaft and one handed natural articulation up to 45 degrees, Precision Machined Anvil to deliver initial, system wide compression, Wide Proximal to Distal Jaw Aperture (proximal 8mm , Distal 22mm) , 3 Point Gap Control for alignment and calibration throughout the 60 mm staple line, Knife direction/reverse control to discontinue the firing and return the knife, interchangeable 6 row cartridge options of white, blue ,gold, green and black, all fits down to 12mm trocar sleeve, 340 mm shaft length
53	(Compact Shaft Length) - Endoscopic linear cutter having 60mm staple line, 57 mm cut line, with single handed natural articulation (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 4- stroke firing with stroke indicator, 360 degree rotation, 280 mm shaft length.
54	Endoscopic linear cutter cartridge having 60mm staple line, White cartridge having closed staple height of 1.0mm compatible with Endoscopic linear cutter 60mm.
55	Endoscopic linear cutter cartridge having 60mm staple line, Blue cartridge having closed staple height of 1.5mm,compatible with Endoscopic linear cutter 60mm.
56	Endoscopic linear cutter cartridge having 60mm staple line, Gold cartridge having closed staple height 1.8 mm, compatible with Endoscopic linear cutter 60mm.
57	Endoscopic linear cutter cartridge having 60mm staple line, Green cartridge having closed staple height of 2mm compatible with Endoscopic linear cutter 60mm.
58	Multifire curved cutter stapler for simultaneous staple and cutting with unique curved head design, 4 curvilinear rows of staple with single cut line, intermediate locking position, 40 mm staple line in width of 30mm space, cutting blade in cartridge.
59	Optically guided bladeless trocar 5mm with bilateral tissue separators, optical tip to eliminate blind entry, clear ribbed cannula to enhance abdominal wall retention, recessed stopcock valve and funnel shaped housing, 75mm length.

60	Optically guided bladeless trocar 11mm with bilateral tissue separators, length 100mm, optical tip to eliminate blind entry, clear ribbed cannula, recessed stopcock valve, funnel shaped housing, duckbill secondary seal, integrated universal seal that eliminates the use of reducer.
61	Sleeve for Bladeless trocar 5mm with bilateral tissue separators, clear ribbed cannula to enhance abdominal wall retention.
62	Titanium ligating clips small
63	Titanium ligating clips medium
64	Titanium ligating clips medium large
65	Titanium ligating clips large
66	Endoscopic rotating clip applier 5 mm with 15 preloaded titanium clips of Medium Large Size (LT300) with shaft length of 33mm & Closed clip length of 8.8mm.
67	Endoscopic rotating clip applier 10 mm with 20 preloaded titanium clips -Medium-Large size(LT300)
68	Endoscopic rotating clip applier 12mm with 20 preloaded titanium clips - large size(LT400)
69	Straight atraumatic 5mm grasper, ergonomically designed with 360° rotation with ratchet mechanism having 34cm length for laparoscopic surgery
70	Curved 5mm scissor, ergonomically designed with 360° rotation with universal monopolar attachment having 34cm length for laparoscopic surgery
71	Curved 5mm dissector, ergonomically designed with 360° rotation with universal monopolar attachment having 34cm length for laparoscopic surgery
72	5mm babcock extractor, ergonomically designed with 360° rotation with ratchet mechanism having 34cm length for laparoscopic surgery
73	laproscopic reusable clip applicator for medium clips LT200
74	laproscopic reusable clip applicator for medium/large clips LT300
75	laproscopic reusable clip applicator for large clips LT400
76	Endoscopic linear cutter having 60mm staple line, 57 mm cut line, with single handed natural articulation (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 4- stroke firing with stroke indicator, 360 degree rotation.440 mm shaft length – (Long Length).
77	Battery Operated 60mm Articulating Endo Cutter with a disposable battery pack, for enhanced distal tip stability while firing, 360 degree rotation shaft and one handed natural articulation up to 45 degrees, Precision Machined Anvil to deliver initial, system wide compression, Wide Proximal to Distal Jaw Aperture (proximal 8mm , Distal 22mm) , 3 Point Gap Control for alignment and calibration throughout the 60 mm staple line, Knife direction/reverse control to discontinue the firing and return the knife, interchangeable 6 row cartridge options of white, blue ,gold, green and black, all fits down to 12mm trocar sleeve, 440 mm shaft length
78	Self-righting reusable needle holder conducting auto-correction for needle holding angles 5mm diameter with palm type handle
79	5mm Lap Hand Activated Curved Coagulating Shears capable of sealing blood vessels up to 5mm in diameter, 36& 45 cm long, Ergonomic Grip.
80	5mm Lap Hand Activated Curved Coagulating Shears capable of sealing blood vessels up to 5mm in diameter, 36 cm long, Ergonomic Grip with Adaptive Tissue Technology.
81	9cm shaft, curved, tapered tip for precise dissection, seals 5 mm vessels, as well as lymphatic with 16 mm active blade & 240- degree activation, triggers support multiple hand positions with and without adaptive Tissue Technology.

82	17cm shaft, curved, tapered tip for precise dissection, seals 5 mm vessels, as well as lymphatic with 16 mm active blade & 240- degree activation, triggers support multiple hand positions.
83	Transducer/hand piece for ultrasonically activated cutting and coagulating device, compatible with open & Lap Ultrasonic shears.
84	5mm Lap Hand Activated Curved Coagulating Shears capable of sealing blood vessels up to 5mm in diameter, 23 cm long, Ergonomic Grip.
85	5mm Lap Hand & Foot Activated with curved & round tip (5&3mm tip width) capable of sealing blood vessels up to 7mm in diameter, simultaneously cut and coagulate tissues, 35 cm long with 110 degree Articulation (55 degree each side) with 360 degree rotation.
86	5mm Lap Hand Activated Curved Coagulating Shears capable of sealing blood vessels up to 7mm in diameter, 36 cm long, Ergonomic Grip with Adaptive Tissue Technology. .
87	5mm open surgery Hand & Foot Activated with round tip & Curved (5mm & 3mm tip width) capable of sealing blood vessels up to 7mm in diameter, simultaneously cut and coagulate tissues, 25 cm long
88	Handle with option of Hand controlled Suction, Irrigation and using Monopolar Energy
89	Insert with Suction Irrigation cannula and having Monopolar Hook, 34 cm shaft length
90	Insert with Suction Irrigation cannula and having Monopolar L Hook, 34 cm shaft length
91	Insert with Suction Irrigation Cannula and having Monopolar Spatula , 34 cm shaft Length
92	LINEAR CUTTER CARTRIDGE 55mm compatible with linear cutter 55mm – <b>Blue</b> cartridge having closed staple height of 1.5 mm with tissue retaining button.
93	LINEAR CUTTER CARTRIDGE 55mm compatible with linear cutter 55mm– <b>White</b> cartridge having closed staple height of 1.0 mm with tissue retaining button.
94	Linear Cutter 75mm having cut line 73 mm and staple line 77 mm with CAM mechanism , intermediate locking position, preloaded with Blue cartridge.
95	LINEAR CUTTER CARTRIDGE <b>75mm compatible with linear cutter 75mm– Green</b> cartridge having closed staple height of 2.0mm with tissue retaining button.
96	LINEAR CUTTER CARTRIDGE 75mm compatible with linear cutter 75mm – <b>Gold</b> cartridge having closed staple height of 1.8mm with tissue retaining button.
97	LINEAR CUTTER CARTRIDGE <b>75mm compatible with linear cutter 75mm– Blue</b> cartridge having closed staple height of 1.5 mm with tissue retaining button.
98	Selectable staple height linear cutter 55mm having 6 rows of staple line, having cartridge with selectable closed staple height of (1.5mm/1.8mm/2mm) & integrated new knife blade. Linear cutter gives intermediate locking position for easy tissue manipulation & two sided firing for easy handling, staple line of 61mm & cut line of 58mm
99	Universal linear cutter cartridge 55mm compatible with selectable staple height linear cutter 55mm. option of closed staple height of 1.5mm/1.8mm/2mm in one cartridge only
100	Selectable staple height linear cutter 75mm having 6 rows of staple line, having cartridge with selectable closed staple height of (1.5mm/1.8mm/2mm) & integrated new knife blade. Linear cutter gives intermediate locking position for easy tissue manipulation & two sided firing for easy handling, staple line of 78mm & cut line of 81mm

101	Universal linear cutter cartridge 75mm compatible with selectable staple height linear cutter 75mm.option of closed staple height of 1.5mm/1.8mm/2mm in one cartridge only
102	Reusable linear cutter/stapler handle 55mm
103	Linear cutter knife module compatible with reusable linear cutter handle 55 mm
104	<b>Linear cutter no- knife module</b> compatible with reusable linear cutter handle 55 mm
105	<b>Linear cutter cartridge Blue 55mm</b> compatible with reusable linear cutter handle having close staple height of 1.5mm
106	<b>Linear cutter cartridge Green 55mm</b> compatible with reusable linear cutter handle having close staple height of 2 mm.
107	Curved cutter stapler reload cartridge –Blue with closed staple height 1.44mm
108	Hemorrhoidal stapler kit consists of <b>33mm Hemorrhoidal Circular stapler</b> (with <b>fixed anvil</b> , adjustable <b>closed staple height from 0.75 mm – 1.5 mm</b> , staple <b>open leg length of 5.5 mm</b> ), Suture Threader, Circular Anal Dilator, Purse-String Suture Anoscope, Suture for Purse String.
109	LX107 liga clip applier for small clips with color-coded handle and lateral/transverse groover corresponding to the clip.
110	LX207 liga clip applier for medium clips with color-coded handle and lateral/transverse groover corresponding to the clip.
111	LX307 liga clip applier for medium large clips with color-coded handle and lateral/transverse groover corresponding to the clip.
112	(Short Shaft Length) – <b>Endoscopic linear cutter having 45mm staple line</b> , 42 mm cut line with single handed <b>natural articulation</b> (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 3+1-stroke firing with stroke indicator, 360 degree rotation,280 mm shaft length
113	(Normal Shaft Length) – <b>Endoscopic linear cutter having 45mm staple line</b> , 42 mm cut line with single handed <b>natural articulation</b> (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 3+1- stroke firing with stroke indicator, 360 degree rotation, 340 mm shaft length.
114	Endoscopic linear cutter cartridge <b>having 45mm staple line</b> ,White cartridge having closed staple height of 1.0mm compatible with Endoscopic linear cutter 45mm.
115	Endoscopic linear cutter cartridge <b>having 45mm staple line</b> ,Blue cartridge having closed staple height of 1.5mm,compatible with Endoscopic linear cutter 45mm.
116	Endoscopic linear cutter cartridge <b>having 45mm staple line</b> , Gold cartridge having closed staple height 1.8 mm, compatible with Endoscopic linear cutter 45mm.
117	Endoscopic linear cutter cartridge <b>having 45mm staple line</b> , Green cartridge having closed staple height of 2mm compatible with Endoscopic linear cutter 45mm.
118	Endoscopic Linear Cutter <b>having 35mm staple line</b> , 32mm cut line, 45 degree bilateral articulation on the instrument (stapler) with 360 degree rotation, along with the cartridges blue (Closed Staple Height 1.5mm) preloaded.
119	Endoscopic Linear Cutter Cartridge 35 mm White cartridge having closed staple height of 1.0mm, compatible with Endoscopic articulating linear cutter 35mm.

120	Endoscopic Linear Cutter Cartridge 35mm Blue cartridge having closed staple height of 1.5mm, compatible with Endoscopic articulating linear cutter 35mm.
121	Endoscopic Linear Cutter <b>having 45mm staple line</b> , 42mm cut line, 45 degree <b>bilateral articulation on the instrument (stapler)</b> with 360 degree rotation.
122	Endoscopic Linear Cutter Cartridge having 45mm staple line, 42 mm cut line, White cartridge having closed staple height of 1.0mm. compatible with Endoscopic articulating linear cutter 45mm.
123	Endoscopic Linear Cutter Cartridge having 45mm staple line, 42 mm cut line, Blue cartridge having closed staple height of 1.5mm, compatible with Endoscopic articulating linear cutter 45mm.
124	Endoscopic linear cutter having 60mm staple line, 57 mm cut line, with single handed <b>natural articulation</b> (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 4- stroke firing with stroke indicator, 360 degree rotation. 340 mm shaft length – (Normal Length).
125	Battery Operated 60mm Articulating Endo Cutter with a disposable battery pack, for enhanced distal tip stability while firing, 360 degree rotation shaft and one handed natural articulation up to 45 degrees, Precision Machined Anvil to deliver initial, system wide compression, Wide Proximal to Distal Jaw Aperture (proximal 8mm , Distal 22mm) , 3 Point Gap Control for alignment and calibration throughout the 60 mm staple line, Knife direction/reverse control to discontinue the firing and return the knife, interchangeable 6 row cartridge options of white, blue ,gold, green and black, all fits down to 12mm trocar sleeve, 340 mm shaft length
126	(Compact Shaft Length) - Endoscopic linear cutter having 60mm staple line, 57 mm cut line, with single handed <b>natural articulation</b> (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 4- stroke firing with stroke indicator, 360 degree rotation, 280 mm shaft length.
127	Endoscopic linear cutter cartridge having 60mm staple line, White cartridge having closed staple height of 1.0mm compatible with Endoscopic linear cutter 60mm.
128	Optically guided bladeless trocar <b>11mm</b> with bilateral tissue separators, length 100mm, optical tip to eliminate blind entry, clear ribbed cannula, recessed stopcock valve, funnel shaped housing, duckbill secondary seal, integrated <b>universal seal that eliminates the use of reducer</b> .
129	Optically guided bladeless trocar <b>12mm</b> with bilateral tissue separators, optical tip to eliminate blind entry, clear ribbed cannula to enhance abdominal wall retention, recessed stopcock valve, funnel shaped housing, duckbill secondary seal, integrated <b>universal seal that eliminates the use of reducer, 100mm length</b> .
130	Optically guided bladeless trocar <b>12mm with Optiview Technology</b> , having wiping, wicking and absorbing technology to reduce smudging during camera reinsertion. With bilateral tissue separators, optical tip to eliminate blind entry, clear ribbed cannula, recessed stopcock valve, funnel shaped housing, duckbill secondary seal, integrated <b>universal seal that eliminates the use of reducer , 100mm length</b>

131	<b>Sleeve of optically guided bladeless trocar compatible with 5mm</b> with bilateral tissue separators, optical tip to eliminate blind entry, clear ribbed cannula to enhance abdominal wall retention, recessed stopcock valve and funnel shaped housing, 75mm length.
132	Sleeve of optically guided bladeless trocar compatible with optically guided bladeless trocar <b>11mm</b> with bilateral tissue separators, optical tip to eliminate blind entry, clear ribbed cannula to enhance abdominal wall retention, recessed stopcock valve, funnel shaped housing, duckbill secondary seal, integrated <b>universal seal that eliminates the use of reducer, 100mm length.</b>
133	Sleeve of optically guided bladeless trocar compatible with optically guided bladeless trocar <b>12mm</b> with bilateral tissue separators, optical tip to eliminate blind entry, clear ribbed cannula to enhance abdominal wall retention, recessed stopcock valve, funnel shaped housing, duckbill secondary seal, integrated <b>universal seal that eliminates the use of reducer, 100mm length.</b>
134	Sleeve of Optically guided bladeless trocar <b>12mm with Optiview Technology</b> , having wiping, wicking and absorbing technology to reduce smudging during camera reinsertion. With bilateral tissue separators, optical tip to eliminate blind entry, clear ribbed cannula, recessed stopcock valve, funnel shaped housing, duckbill secondary seal, integrated <b>universal seal that eliminates the use of reducer , 100 mm length</b>
135	Bladeless trocar 11mm with bilateral tissue separators, integrated universal seal that eliminates the use of reducer, clear ribbed cannula to enhance abdominal wall retention ,
136	Bladeless trocar 12mm with bilateral tissue separators, integrated universal seal that eliminates the use of reducer, clear ribbed cannula to enhance abdominal wall retention.
137	Bladeless trocar 5mm with bilateral tissue separators, clear ribbed cannula to enhance abdominal wall retention.
138	<b>Sleeve for Bladeless trocar 11mm</b> with bilateral tissue separators, integrated universal seal that eliminates the use of reducer, clear ribbed cannula to enhance abdominal wall retention ,
139	<b>Sleeve for Bladeless trocar 12mm</b> with bilateral tissue separators, integrated universal seal that eliminates the use of reducer, clear ribbed cannula to enhance abdominal wall retention.
140	Straight atraumatic 5mm grasper, ergonomically designed with 360° rotation with ratchet mechanism having 34cm length for laparoscopic surgery
141	Curved 5mm scissor, ergonomically designed with 360° rotation with universal monopolar attachment having 34cm length for laparoscopic surgery
142	Curved 5mm dissector, ergonomically designed with 360° rotation with universal monopolar attachment having 34cm length for laparoscopic surgery
143	5mm babcock extractor, ergonomically designed with 360° rotation with ratchet mechanism having 34cm length for laparoscopic surgery
144	Endoscopic linear cutter having 60mm staple line, 57 mm cut line, with single handed <b>natural articulation</b> (options of 15 degree, 30 degree, 45 degree on either side), 3-point Gap control, 4- stroke firing with stroke indicator, 360 degree rotation.440 mm shaft length – (Long Length).



145	Battery Operated 60mm Articulating Endo Cutter with a disposable battery pack, for enhanced distal tip stability while firing, 360 degree rotation shaft and one handed natural articulation up to 45 degrees, Precision Machined Anvil to deliver initial, system wide compression, Wide Proximal to Distal Jaw Aperture (proximal 8mm , Distal 22mm) , 3 Point Gap Control for alignment and calibration throughout the 60 mm staple line, Knife direction/reverse control to discontinue the firing and return the knife, interchangeable 6 row cartridge options of white, blue ,gold, green and black, all fits down to 12mm trocar sleeve, 440 mm shaft length
146	Battery Operated 35 mm Articulating Endo Cutter with a disposable battery pack, for enhanced distal tip stability while firing, 360 degree rotation shaft and manual articulation up to 51 degrees either side and curved tip for better visuability.
147	White reload with closed staple height of 1.0mm compatible with Powered Vascular Gun