

Technical Corrigendum

Tender Enquiry No. 24/RT/MEC/217/2020-Rish (Admn)

S.No	Name of the Equipment with Tender Specification	Corrigendum/Amendment
1	ACLS Training system with wireless control and patient monitor	
	Point No 13. The Training system should have ability to get connected to an optional simulated patient monitor to display various vital signs with simulated 12 leads ECG display	After amendment read as: The Training system should have ability to get connected to an simulated patient monitor to display various vital signs with simulated 12 leads ECG display
2	Item Code 02: Advance CPR Training System with Feedback	
	Technical Specification: It should be upgradable to integrate the optional trauma/rescue limbs and airway head	After amendment read as: It should be supplied with Trauma, rescue limbs and airway head