

Corrigendum-2
In tender document
Tender Enquiry No. 24/OT/328(III)/2018-Rish(Admn)

Dated: 12-05-2018

After consideration by Technical Specification Committee and corrigendum 1 issued on dated 14-05-2018 following modification (deletions/additions/replacements) additions for Tender Enquiry 24/OT/328(III)/2018-Rish(Admn) has been made.

Point Srl no. 05(f) at page no. 18: -

For: - SPO2 approx 10 to 10% accuracy +1%

Read as: - Masimo SET SpO2: 10 to 100% accuracy \pm 1% with SpO2. NIBP

Point Srl no. 13 at page no. 18: -

For: - Trend display (Numerical & graphic) from 24 hours. Facility to zoom in up to 1 min. The trend data should not be lost on switching off the monitor.

Read as: - Trend display (Numerical & graphic) from 220 hours. Should also have full disclosure of waveforms for minimum 48 hours. Minimum 200 events recall with 3 wave forms snapshots.

Point Srl no. 17 at page no. 18: -

For: - Power requirement: 220V/50 Hz (with adapter) and integral rechargeable batteries with (autonomy at least 2 hours, automatic recharge)

Read as: - Power requirement: 220V/50 Hz (with adapter) and integral rechargeable batteries with (continuous backup of least 4-6 hours, automatic recharge)

Following point is being added: -

1. Monitor should be upgradable to monitor non-invasive haemoglobin, dual IBP and micro-stream ETCO2. Price of these parameters to be quoted separately.