

Corrigendum
In tender document
Tender Enquiry No. 24/Venti/386/2018-Rish(Admn)

Dated: 14-08-2018

As per schedule, Pre- Bid meeting of "Rate Contract for Advance Intensive Care Ventilator" was held on 17-07-2018 at 03.00 PM, in the tender opening room.

After consideration by Technical Specification Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/Venti/386/2018-Rish(Admn)" has been made.

Point Srl no. 06 (a) at page no. 17: -

For: - Tidal Volume : 2ml to 4000 ml.

Read as: -

Tidal Volume : 2ml to 2000 ml.