

Corrigendum

Tender enquiry no 24/Urology/200 W Thulium/518/2019-Rish(Admn)

Date: 11/09/2019

In reference to Pre Bid meeting date 27-08-2019 of Tender documents for procurement of 200W Thulium Surgical Laser for Urology AIIMS Rishikesh, Tender Office, following modification for tender enquiry no 24/Urology/200 W Thulium/518/2019-Rish(Admn) has been made from the Department.

Point Sr. No. 2

For: - Maximum output of Thulium Laser should be: 200W

Read As: - **Maximum output of Thulium Laser should be: 200W & to be covered under warranty & CMC**

Point Sr. No. 3

For: - Mode of Operation in Laser: Continuous & Pulse and Burst Mode

Read As :- Mode of Operation in Laser: Continuous & Pulse Mode

Point Sr. No. 7

For: - Fiber Holding stand with adjustable position

Read As:- This point should be deleted.

Point Sr. No. 10

For: - Software upgradation facility should be there

Read As :- This point should be deleted.

Point Sr. No. 19.

For: - **Thulium laser should be operated with minimum of 1 W with 200 um laser fiber.**

Read As :- Thulium laser should be operated with minimum of 5 W with 200/270 um micron fiber.

Point Sr. No. 22

For: - Repetition frequency in pulse mode: 0.2 to 200 Hz.

Read As :- This point should be deleted.

Point Sr. No. 24

For :- The machine should be supplied with compatible complete morcellator with morcelloscope

Read As: - The machine should be supplied with compatible complete morcellator with morcelloscope of repute make.

Point Sr. No. 25

For: - F. Laser Ribres 600um – 10 Nos.
G. Laser Ribres 200 um – 5 Nos. for upper tract pathologies

Read As: - F. Laser Fibres 550/600um – 30 Nos.
G. Laser Fibres 200/272um um – 20 Nos. for upper tract pathologies
H. Laser Fibres 365/400um – 10 Nos.

Point Sr. No. 28

For: - At least 3 maintenance visit to be made by service engineer per year in warranty period. this is apart
from the repair / breakdown call.

Read As: - At least 4 maintenance visit to be made by service engineer per year in warranty period. this is apart from the repair / breakdown call.

Point Sr. No. 32

For: - Backup in case of offsite repair of equipment.

Read As: - Backup equipment in case of any damage including physical failure of equipment to be provided within 24 hrs from mail receiving time.

Addendum

Point Sr. No. 33

For: - Point to be added

Read As: - 1. Service Manual to be provided with machine.

2. Following items from reputed company should be added: -

- Laser Resectoscope – 1 Set
- Laser Cystoscope - 1 Set
- 30 degree Telescope – 2 Nos.

Point Sr. No. 34

For: - Point to be added

Read As: - 5 Year warranty & CMC & CMC for 6th to 10th year.