

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, RISHIKESH

Tender No. 24/FTIR/Biochemistry/533/2019-RISH (ADMN)

Corrigendum

As per schedule, Pre-bid meeting for procurement of Fourier-transform infrared (FTIR) Spectrometer held on 12/12/2019 at Tender Office following Changes recommended for procurement of Department of Biochemistry.

After consideration by Store Purchase Committee following modification (Deletion/ Addition/ Replacement) for tender for 24/FTIR/Biochemistry/533/2019-RISH (ADMN) has been made:

Sr No.	Original	Should be read as
Point No. 19: Page no. 15	System should be certified by CE, ISO and NIST.	System Should Be certified by ISO and BIS/ USFDA/ NIST/ European CE or one equivalent. Instrument will be used for diagnostic services so it must be IVD approved.
Point No. 37; Page No. 16	Also 1 unit of each accessory and consumables should be supplied as FOC along with instrument.	Also 1 unit of each accessory and consumables (Pellet Holder, KBr Powder, and Liquid Cell with KBr Windows, Agate Mortar and Pestle, and 15-ton Hydraulic Press with KBr Dye) should be supplied as FOC along with instrument.
Point No. 36; Page No. 16	The accessories and consumables required for functionality of instrument rates to be freeze for 3 years extendable to 2 years.	The accessories and consumables required for functionality of instrument (Pellet Holder, KBr Powder, and Liquid Cell with KBr Windows, Agate Mortar and Pestle, and 15-ton Hydraulic Press with KBr Dye) rates to be freeze for 3 years extendable to 2 years.