



अखिल भारतीय आर्युविज्ञान संस्थान ऋषिकेश-249203
ALL INDIA INSTITUTE OF MEDICAL SCIENCES Rishikesh Uttarakhand
Procurement of Department of Physiology
Tender No. 24/Physio/NCV/EMG/570/2020-Rish (Admin)

Corrigendum

In reference to mail dated on 28/05/2020 from **Krishan Yogi** krishan_yogi@nihonkohden.in
Subject: Amendment request in Tender 24/Physio/NCV/EMG/570/2020-Rish(Admn) for
Purchase Portable NCV/EMG recording system

Tender Enquiry No. F. No. 24/Physio/NCV/EMG/570/2020-Rish (Admin)
Following corrigendum may be considered

1. **Point 1: Point under system hardware:** The amplifier should have USB powered and enabling the amplifier to be connected to any PC/ Laptop and transfer data through same USB.
 - a. Read as – The amplifier/Main unit should have USB/AC powered and the amplifier/Main unit to be connected to any PC/Laptop and transfer data through USB/Ethernet technology.
2. **Point 2: Point under system hardware:** The system should have wireless/wired control panel for easy operation.
 - a. Read as – The system should have wireless /wired OR Built-in control panel on Main Unit for easy operation.
3. **Point 3: Point under EMG/NCV/EP Software:** Tremor analysis
 - a. Point to be considered deleted.
4. **Point 4: Point under EMG/NCV/EP Software:** Dedicated control panel can function via Bluetooth or USB interface.
 - a. Read as- Dedicated /Built-in Control panel can function via Bluetooth or USB interface.
5. **Point 5: Point supplies with EMG system:** Gold plated cup electrode – 10 pcs.
 - a. Read as – Gold/Silver plated cup electrode – 10 pcs.