

Corrigendum
In tender document
Tender Enquiry No. 24/Transfusion/308(II)/2018-Rish(Admn)”

Dated: 07-04-2018

As per schedule, Pre- Bid meeting of “Tender for Blood Bank Monitoring Surveillance Software System for Department of Transfusion Medicine”. was held on 14.03.2018 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/Transfusion/308(II)/2018-Rish(Admn)” has been made.

Query 1: Sole manufacturers must add a certificate that they are the sole manufacturer of the item for which they are quoting in this tender enquiry & item is/are their proprietary item in India. The rate certificate is also required from the sole manufacturers that the rates quoted are the same as they quote to other state/Centre Govt./reputed private organization and DGS & D rate for the similar item(s) and these are not higher than those quoted by them- Need more clarity

Comment: This specification is self-explanatory and does not require any further clarification.

Query 2: Please submit two performance certificate from your two different customers to whom you have supplied such type of equipment in previous 3 years- what documents required/recommendation letter or PO (JKT can submit only two references).

Comment: The supplier will require to submit purchase order (P.O.) from the customer.

Query 3: Hardware quantity not specified.

Comment: The hardware quantity required is as under:

| Computer Hardware Requirement | | | | | | |
|--------------------------------------|--------------------------|-----------------|---|------------|------------|---|
| S. No. | Particulars | Quantity | Processor | RAM | HDD | Operating System |
| 1 | Server | 1 | Intel Core i7/Xeon, 3.10 GHz or Higher supported for minimum win 2008 server, 2 HDD of 500 GB | Min 16 GB | 500 GB | Window server 2008 or above, sql server 2012 (Pro or Express) |
| 2 | Desktop | 8 Units | Pentium (Dual core) or higher | Min 2 GB | 160 GB | Window XP (SP2)/ Window 7/Window 8/ Window10 |
| | | | Model | | | |
| 3 | Printer (Label) | 2 Nos. | | | | |
| 4 | Printer (Bill Receipt) | 2 Nos. | Laser | | | |
| 5 | Printer- General Reports | 1 Nos. | A4 laser Printer | | | |

| | | | | | | |
|----|-------------------------------------|--------|----------------------|--|--|--|
| 6 | Printer (Records Register) | 1 Nos. | A3 laser printer | | | |
| 7 | Bar-Code Scanner | 8 Nos. | USB Bar Code Scanner | | | |
| 8 | Switch (Network Connectivity) | 1 Nos. | D-link 16-Port | | | |
| 9 | Network Cable- UTP (Cat 6) | 1 Box | D-link | | | |
| 10 | Finger printer Scanner | 2 Nos. | | | | |
| 11 | IRIS Scanner | 1 Nos. | | | | |
| 12 | TABS for inhouse donor registration | 3 Nos. | | | | |
| 13 | Laptop | 1 Nos. | | | | |
| 14 | Smart Card Printer | 1 Nos. | | | | |

Note: - Required continuous internet connectivity on server computer required Quick Heal Antivirus.

Query 4: Project duration not specified.

Comment: The project duration is 1 month.

Query 5: Smart card integration- who will provide the smart card either vendor or customer i.e. AIIMS.

Comment: The smart card will be provided by the vendor.

Query 6: Payment terms not cleared.

Comment: The payment terms are mentioned at points srl. 05 A & B at page no. 08 & 09.