

Corrigendum- Changes recommended for procurement of NO Generator for Department of CTVS.

Point number	original	Should be read as
06	The tendering firm should undertake to supply custom-made B type regulators which can be fitted to the preexisting No cylinders available presently in CTVS Deptt. If need be, the firms quoting may inspect the available NO cylinders to supply the B type regulators.	The tendering firm should undertake to supply custom made compatible regulators which can be fitted to the NO cylinders and price for refilling should be fixed for next two years on pick up & drop services at Rishikesh.
09	Warranty – There would be warranty for 5 years. If any part of the equipment is not covered under warranty, the vendor must provide advance information. The cost of disposable accessories (if any) should be quoted separately. This quoted price will be fixed for 5 years period.	Warranty- there should be warranty for 05 year. If any part of the equipment is not covered under warranty, the vendor must provide advance information. The cost of disposable accessories (if any) should be quoted separately. This quoted price will be fixed for 02 -year period.
Additional Point – The whole unit should include		
1.		Main unit
2.		Trolley
3.		Calibration cylinder (s)
4.		Nitric oxide cylinder 10 lit. capacity with 900 ppm NO in N2 Qty. 10 nos.
5.		Patient circuit for neonatal, paediatric & adult Qty. 10 nos. each.
6.		All consumables, maintenance kits, accessories required to be changed to be changed periodically must be quoted with separate price, which shall remain frozen for next 2 year.

12/3/2020

 31/03/2020

 06/03/2020