

AIMS/RATH/RS&S/2020/422

CORRIGENDUM-3 FOR VACUTAINER

S.No	Items Name	To be read as	Current Specification	To be read as	Current Pack Size	To be read as
1	Blood collection tube evacuated (Blue)	No change	Sodium Citrate for coagulation	<p>a) It should be made of clear latex free polyethylene terephthalate (PET) with preferred spill proof non screw Blue closure, 13 mm x 75mm Vol. 2.7 ml.</p> <p>b) It should have Tube-in-Tube Technology for Full Draw Volume that reduces dead space thereby minimize platelet activation and provide accurate APPT result</p> <p>c) Removed</p>	100/pkt.	100/Pkt
2	Blood collection tube evacuated	Blood collection tube Paediatric	EDTA-Paediatric tube	<p>a) Paediatric Blood Collection tubes with Spray dried K2EDTA</p> <p>b) It should be Capped and made of clear latex free polyethylene terephthalate (PET)</p> <p>c) Fill volume should be 250 – 500µl with marking.</p> <p>d) It should have protruded edge to collect blood in easy flow.</p>	100/pkt.	100/pkt.

CORRIGENDUM-3 FOR VACUTAINER

3	Blood collection tube evacuated	No change	Oxalate/fluoride tube Plain/Non additive	<p>a) It should be spray coated with Sodium Fluoride (3mg.), Na₂EDTA (6mg.) / Pot. Oxalate made of clear latex polyethylene terephthalate (PET) with preferred spill proof non screw Grey closure 13 mm x 75 mm, Vol. 2.0 ml /4ml</p> <p>b) Removed</p>	100/pkt	100/pkt
4	Blood collection tube evacuated	No change	Plain/Non additive	<p>a) It should have Silica-clot activator/silicon coated made of clear latex polyethylene terephthalate (PET) with preferred spill proof non screw Red closure 13 mm x 75 mm, Vol. 4.0 ml.</p> <p>b) Removed</p>	100/pkt	100/pkt

CORRIGENDUM-3 FOR VACUTAINER

5	Blood collection tube evacuated	No change	EDTA	<p>a) It should have spray dried K2EDTA made of clear latex free polyethylene terephthalate (PET) with preferred spill proof non screw Lavender Cap. Size 13 x 75 mm with 3.0 ml (5.4 mg), volume capacity. b) Removed</p>	100/pkt	100/pkt
6	Blood collection tube evacuated	No change	Green Top	<p>a) It should be spray coated with Lithium or Sodium Heparin made of clear latex polyethylene terephthalate(PET) with preferred spill proof non screw Green closure 13 mm x 75 mm, Vol. 4.0 ml. b) Removed</p>	100/pkt	100/pkt
7	Needle for evacuated Blood collection tube	No change	--	<p>a) Needle should be of size 22G. b) Should be compatible with all types of vacutainers c) Should preferably be from the same manufacturer as tubes to ensure compatibility.</p>	100/pkt	100/pkt
8	Needle holder for evacuated Blood collection tube		No specs given	Should preferably be from the same manufacturer as tubes and needles for compatibility	01 No.	50-100/Pkt

CORRIGENDUM-3 FOR VACUTAINER

9	Blood collection tube evacuated	No change	Sod citrate paediatric tube	<p>a) It should be made of clear latex free polyethylene terephthalate (PET) with preferred spill proof closure, 13 mm x 75mm Vol. 1.8 ml.</p> <p>b) It should have Tube-in-Tube Technology for Full Draw</p> <p>Volume that reduces headspace thereby minimize platelet activation and provide accurate APPT result</p> <p>c) Removed</p>	100/Pkt	100/Pkt
10	Blood collection tube evacuated	Blood collection tube Paediatric	Red Paediatric	<p>a) It should be Capped and Silica-clot activator/silicon coated made of clear latex polyethylene terephthalate(PET)</p> <p>b) Fill volume should be 400 – 600µl with marking.</p> <p>c) It should have protruded edge to collect blood in easy flow.</p>	100/Pkt.	100/Pkt

CORRIGENDUM-3 FOR VACUTAINER

Common Points for all items	
1	Batch wise sterility/pyrogenicity/Toxicity certificate should be given with sample and supplies
2	Adequate combustion data to prove that it is safe for the environment upon incineration.
3	Firm should provide training to technicians/nurses regularly to ensure supplied products used as per recommendations for best outcomes. Clinical Services to be documented by company
4	BIS(EQUIVALENT INDIAN STANDARD) OR EU-CE/USFDA OR ISO6710 compliant
5	Proof of manufacturing for 10 years
6	The companies will have to provide a sample of each item quoted, for demonstration. The samples may be rejected upon demonstration if found unsuitable.