## Corrigendum

## In Tender document

Tender Enquiry no. 24/RC/Suture/564/2020-Rish(Admn)

Dated: 22/04/2020

As per schedule, pre-bid Meeting of "RC for sutures" was held on 22/04/2020 at 10.00 AM in Department of General Surgery.

After consideration by the Suture Committee, the following modifications (addition/replacement) for tender enquiry no.24/RC/Suture/564/2020-Rish(Admn) have been made on technical specifications:

Point No 1.

For: Serial No. 149,150,151 "Synthetic Absorbable Coated Unidirectional Barbed Dyed glycolide + dioxanone + Trimethylene carbonate"

<u>Read as</u>: "Synthetic Absorbable Coated Unidirectional Barbed Dyed glycolide + dioxanone + Trimethylene carbonate /Poliglecaprone 25/PGA-PCL Barbed"

## Point No 2.

For: Serial No 202-203 "Polyglicolic Barbed"

Read as: "Polyglicolic Barbed/Polydiaxanone Barbed"

## Note: Addendum of new sutures

Corrigendum to the tender: Addition of following sutures with annual demand with the following serial numbers 227-229.

227. Non absorbable blue monofilament polypropylene rhenium tungsten alloy, 3/8 circle TP point Double needle, 9.3mm, suture length 60cm, size 7-0. 72 Pieces

228. Non absorbable blue monofilament polypropylene rhenium tungsten alloy, 3/8 circle TP point Double needle, 8mm, suture length 60cm, size 8-0. 72 Pieces

229. Non absorbable blue monofilament polypropylene rhenium tungsten alloy, 3/8 circle TP point Double needle, 6.5mm, suture length 60cm, size 8-0. 72 Pieces

The above added items should be quoted in pdf format along with BoQ (Bill of

Quantity)

अक्षय कपूर/Dr. Anshay Kapoor

सहोयक-आचार्य, प्लास्टिक चिकित्सा विभाग Assistant Professor of Burns and Plastic Surgen एम्स, ऋषिदेश/AIIMS, Rishikanh डॉ. एनोनो यहोशु / Dr. Enono Yhoshu सहायक आचार्य /Assistant Professor बाल शल्य चिकित्सा विभाग Department of Paediatric Surgery

ऋषिकेश/AIIMS, Rishikesh