

Tender Enquiry No. F. No. 24/Crash Cart/500/2019-RISH (ADMN)

Corrigendum (24/crash cart/500/2019/Rishi admin)

1. Overall size shall be more than 1100 (height) x 800 mm (width) x 700mm (depth).

To be read as: Overall size shall be L690-700 x W565-600 x H1010-1100mm
Approximately.

2. The crash cart should be made of 25.4 mm 18G stainless steel tubular framework.

To be read as: Crash cart should be made of ABS.

3. Should have Epoxy/Anti - microbial powder paint inside and out.

To be read as: DELETED

4. Should have dual push handles on either side.

To be read as: Should have push handles on both sides

5. Should have S.S shelves, six colored removable drawers. Upper 2 drawers shall have 25 to 30 compartments each for storing drugs.

To be read as: Should have 5 ABS removable drawers. All drawers shall have provision for 25 compartments each.

6. Crash cart should contain 3 small depth 2 medium depth and 1 large depth shelves.

To be read as: Crash cart should contain 2 small depth 2 medium depth and 1 large depth shelves.

7. Facility to carry ECG monitor, defibrillators, on open areas on top centre.

To be read as: No change

ban

SS

han

Chini

Jan

8. Should have stainless steel saline rod fixed with.

To be read as: No change

9. Two accessory mounting brackets to mount accessories anywhere without the need to pre threaded holes.

To be read as: No change

10. Crash cart should have side storage for keeping cans, storage bins, gloves dispenser, sharp container.

To be read as: No change

11. Crash cart should be mounted on 12.5 cms dia non – rusting swiveling castor wheels. Two having locking arrangements. Wheel should be of 4 to 5 inches diameter for easy rolling and with foot brakes.

To be read as: No change

12. Oxygen cylinders stand epoxy powder coated on one side.

To be read as: Oxygen cylinder stand epoxy powder coated on one side or at the back.

13. Crash cart should have a hard board for performing CPR.

To be read as: No change

14. Crash cart should have central locking system.

To be read as: No change

15. Manufacturer should be NABL certified/ISO/BIS/European CE/USFDA

To be read as: no change

Points to be added:

16. Should have 2 no. Waste Bin and 1 no. Needle disposal container.

17. Should have power strip and power strip holder.

bel

[Signature]

[Signatures]

- 18. Should have PVC top mat
- 19. Should have sliding side shelf
- 20. Should have file cassette
- 21. Should have label cards option for drawers

Overall size shall be ...
 (depth)
 To be read as: Overall size ...
 Approximately.

The crash cart should be ...
 framework
 To be read as: Crash cart should be ...

but *SS* *lean* *crim* *alm*

Should have dual push handles on other side.
 To be read as: Should have push handles on both sides

Should have SS shelves, six colored removable drawers. Upper 2 drawers shall have 25 to 30 compartments each for storing drugs.
 To be read as: Should have 5 ABS removable drawers. All drawers shall have provision for 25 compartments each.

Crash cart should contain 3 small depth 2 medium depth and 1 large depth shelves.
 To be read as: Crash cart should contain 2 small depth 2 medium depth and 1 large depth shelves.

Facility to carry ECG monitor, defibrillators, on open areas on top centre.
 To be read as: No change

but *SS* *lean* *crim* *alm*